

SINGAPORE STANDARD

Medical electrical equipment

– Part 2-39 : Particular requirements for basic safety and essential performance of peritoneal dialysis equipment

Published by

Enterprise
Singapore

SS IEC 60601-2-39 : 2018

IEC 60601-2-39:2018, IDT
(ICS 11.040.99)

SINGAPORE STANDARD

Medical electrical equipment

– Part 2-39 : Particular requirements for basic safety and essential performance of peritoneal dialysis equipment

All rights reserved. Unless otherwise specified, no part of this Singapore Standard may be reproduced or utilised in any form or by any means, electronic or mechanical, including photocopying and microfilming, without permission in writing from Enterprise Singapore. Request for permission can be sent to: standards@enterprisesg.gov.sg.

© IEC 2018 – All rights reserved
© Enterprise Singapore 2018

ISBN 978-981-48-3507-7

This Singapore Standard was approved by the Biomedical and Health Standards Committee on behalf of the Singapore Standards Council on 27 July 2018.

First published, 2018

The Biomedical and Health Standards Committee, appointed by the Standards Council, consists of the following members:

	Name	Capacity
Chairman	: Dr Yong Chern Chet	<i>Individual Capacity</i>
1st Deputy Chairman	: Ms Wong Woei Jiuang	<i>Health Sciences Authority</i>
2nd Deputy Chairman	: Mr Vincent Cheung	<i>Individual Capacity</i>
3rd Deputy Chairman	: Ms Selina Seah	<i>Changi General Hospital</i>
Advisor	: Ms Jacqueline Monteiro	<i>Individual Capacity</i>
Secretary	: Ms Iris Peng	<i>Singapore Manufacturing Federation – Standards Development Organisation</i>
Members	: Prof Kishore Bhakoo	<i>Singapore Bioimaging Consortium</i>
	Mr Chung Kwong Yuew	<i>Temasek Polytechnic</i>
	Ms Heidi Goh	<i>Singapore Manufacturing Federation (Medical Technology Industry Group)</i>
	Prof James Goh	<i>Biomedical Engineering Society (Singapore)</i>
	Dr Lai Choon Sheen	<i>Eu Yan Sang International Ltd</i>
	Dr Christopher Lam	<i>Health Sciences Authority</i>
	Ms Audrey Lee	<i>Medtronic International Ltd</i>
	Assoc Prof Leo Hwa Liang	<i>National University of Singapore</i>
	Dr Lin Jianhua	<i>TÜV SÜD PSB Pte Ltd</i>
	Dr Leonard Loh	<i>Nanyang Polytechnic</i>
	Assoc Prof Eddie Ng Yin Kwee	<i>Nanyang Technological University</i>
	Dr Ong Siew Hwa	<i>Acumen Research Laboratories Pte Ltd</i>
	Mr Peh Ruey Feng	<i>Advent Access Pte Ltd</i>
	Mr Justin Phoon	<i>Biomedical Research Council</i>
	Dr Padmanabhan Saravanan	<i>Temasek Polytechnic</i>
	Ms Celine Tan	<i>Enterprise Singapore</i>
	Prof Tan Puay Hoon	<i>Singapore Health Services Pte Ltd</i>
	Ms Wang Dan	<i>Biosensors International Group</i>
	Dr Sidney Yee	<i>DxD Hub</i>

The Technical Committee on Medical Devices, appointed by the Biomedical and Health Standards Committee and responsible for the preparation of this standard, consists of representatives from the following organisations:

	Name	Capacity
Chairman	: Prof James Goh	<i>Individual Capacity</i>
Secretary	: Ms Iris Peng	<i>Singapore Manufacturing Federation – Standards Development Organisation</i>
Members	: Mr Chua Chui Khim	<i>Becton Dickinson Medical (S) Pte Ltd</i>
	Dr Christopher Lam	<i>Health Sciences Authority</i>
	Assoc Prof Leo Hwa Liang	<i>Biomedical Engineering Society (Singapore)</i>
	Dr Lim Jing	<i>Osteopore International Pte Ltd</i>
	Ms Iris Tan	<i>Advent Access Pte Ltd</i>

The National Mirror Working Group on IEC SC62D MT20, appointed by the Technical Committee to assist in the preparation of this standard, comprises the following experts who contribute in their individual capacity:

	Name
Convenor	: Dr Christian Gert Bluechel
Secretary	: Mr Kevin Tan
Members	: Mr Peter Haywood
	Dr Ho Teck Tuak
	Mr Watson Ong
	Mr Peh Ruey Feng
	Prof Tong Yen Wah

The organisations in which the experts of the National Mirror Working Group are involved are:

Advent Access Pte Ltd
AWAK Technologies Pte Ltd
IES Biomedical Committee
Kidney Dialysis Foundation
National University of Singapore
Temasek Polytechnic

(blank page)

CONTENTS

NATIONAL FOREWORD	6
FOREWORD	7
INTRODUCTION	10
201.1 Scope, object and related standards	11
201.2 Normative references	12
201.3 Terms and definitions	13
201.4 General requirements	14
201.5 General requirements for testing ME EQUIPMENT	17
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	17
201.7 ME EQUIPMENT identification, marking and documents	17
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	19
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	19
201.10 Protection against unwanted and excessive radiation HAZARDS	19
201.11 Protection against excessive temperatures and other HAZARDS	19
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	21
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	22
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	22
201.15 Construction of ME EQUIPMENT	22
201.16 ME SYSTEMS	23
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	23
202 Electromagnetic disturbances – Requirements and tests	23
208 * General requirements, tests and guidance for ALARM SYSTEMS in MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS.....	24
209 Requirements for environmentally conscious design	25
211 Requirements for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS used in the HOME HEALTHCARE ENVIRONMENT	25
Annexes	26
Annex G (normative) Protection against HAZARDS of ignition of flammable anaesthetic mixtures	27
Annex AA (informative) Particular guidance and rationale.....	28
Bibliography	29
Index of defined terms used in this particular standard	30
Table 201.101 – ESSENTIAL PERFORMANCE requirements.....	14

National Foreword

This Singapore Standard was prepared by the National Mirror Working Group for IEC SC62D MT20 appointed by the Technical Committee on Medical Devices under the direction of the Biomedical and Health Standards Committee.

This Singapore Standard is identical with IEC 60601-2-39:2018, "Medical electrical equipment – Part 2-39: Particular requirements for basic safety and essential performance of peritoneal dialysis equipment" published by the International Electrotechnical Commission.

Attention is drawn to the possibility that some of the elements of this Singapore Standard may be the subject of patent rights. Enterprise Singapore shall not be held responsible for identifying any or all of such patent rights.

NOTE

1. *Singapore Standards (SSs) and Technical References (TRs) are reviewed periodically to keep abreast of technical changes, technological developments and industry practices. The changes are documented through the issue of either amendments or revisions.*
2. *An SS or TR is voluntary in nature except when it is made mandatory by a regulatory authority. It can also be cited in contracts making its application a business necessity. Users are advised to assess and determine whether the SS or TR is suitable for their intended use or purpose. If required, they should refer to the relevant professionals or experts for advice on the use of the document. Enterprise Singapore shall not be liable for any damages whether directly or indirectly suffered by anyone or any organisation as a result of the use of any SS or TR.*
3. *Compliance with a SS or TR does not exempt users from any legal obligations.*

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-39: Particular requirements for basic safety and essential performance of peritoneal dialysis equipment

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-39 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 2007. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) update of the references to IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, of references and requirements to IEC 60601-1-2:2014, of references to IEC 60601-1-6:2010 and IEC 60601-1-

6:2010/AMD1:2013, of references and requirements to IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012 and of references and requirements to IEC 60601-1-11:2015;

- b) editorial improvements;
- c) improvement of the essential performance requirements clause/subclauses;
- d) new requirements for the interruption of the power supply.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/1558/FDIS	62D/1586/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of PERITONEAL DIALYSIS ME EQUIPMENT.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-39: Particular requirements for basic safety and essential performance of peritoneal dialysis equipment

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 Scope

Replacement:

This part of IEC 60601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of PERITONEAL DIALYSIS ME EQUIPMENT as defined in 201.3.208, hereafter referred to as PD EQUIPMENT. It applies to PD EQUIPMENT intended for use either by medical staff or under the supervision of medical experts, including PD EQUIPMENT operated by the PATIENT, regardless of whether the PD EQUIPMENT is used in a hospital or domestic environment.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

This document can also be applied to PD EQUIPMENT used for compensation or alleviation of disease, injury or disability.

These particular requirements do not apply to the DIALYSING SOLUTION, or the DIALYSING SOLUTION CIRCUIT.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for PD EQUIPMENT as defined in 201.3.208.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2:2014, IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012, and IEC 60601-1-11:2015 apply as modified in Clauses 202, 208 and 211. IEC 60601-1-3 and IEC 60601-1-12 do not apply. IEC 60601-1-9:2007 and IEC 60601-1-9:2007/AMD1:2013 do not apply as noted in Clause 209. All other published collateral standards in the IEC 60601-1 series apply as published.

¹ The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 are referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"*Replacement*" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"*Addition*" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"*Amendment*" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.147, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, for example 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the bibliography.

Clause 2 of the general standard applies, except as follows:

Replacement:

IEC 60601-1-2:2014, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests*

IEC 60601-1-6:2010, *Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability*
IEC 60601-1-6:2010/AMD1:2013

IEC 60601-1-8:2006, *Medical electrical equipment – Part 1-8: General requirements for basic safety and essential performance – Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems*
IEC 60601-1-8:2006/AMD1:2012

Addition:

IEC 60601-1-11:2015, *Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*