

**SINGAPORE STANDARD**

## **Medical electrical equipment**

– Part 1-2: General requirements for basic safety and essential performance – Collateral standard:  
Electromagnetic disturbances – Requirements and tests

Published by

**Enterprise**  
**Singapore**

## **SS IEC 60601-1-2 : 2018**

IEC 60601-1-2:2014, IDT

(ICS 11.040.01; 33.100.10; 33.100.20)

---

SINGAPORE STANDARD

### **Medical electrical equipment**

– Part 1-2: General requirements for basic safety and essential performance –

Collateral standard: Electromagnetic disturbances – Requirements and tests

---

All rights reserved. Unless otherwise specified, no part of this Singapore Standard may be reproduced or utilised in any form or by any means, electronic or mechanical, including photocopying and microfilming, without permission in writing from Enterprise Singapore. Request for permission can be sent to: [standards@enterprisesg.gov.sg](mailto:standards@enterprisesg.gov.sg).

© IEC 2014 – All rights reserved

© Enterprise Singapore 2018

ISBN 978-981-48-3527-5

This Singapore Standard was approved on 19 October 2018 by the Biomedical and Health Standards Committee under the purview of the Singapore Standards Council

First published, 2019

The Biomedical and Health Standards Committee, appointed by the Standards Council, consists of the following members:

	<b>Name</b>	<b>Capacity</b>
<b>Chairman</b>	: Dr Yong Chern Chet	<i>Individual Capacity</i>
<b>Deputy Chairmen</b>	: Mr Vincent Cheung Ms Selina Seah Ms Wong Woei Jiuang	<i>Individual Capacity</i> <i>Changi General Hospital</i> <i>Health Sciences Authority</i>
<b>Advisor</b>	: Ms Jacqueline Monteiro	<i>Individual Capacity</i>
<b>Secretary</b>	: Ms Iris Peng	<i>Singapore Manufacturing Federation – Standards Development Organisation</i>
<b>Members</b>	: Mr Alec Chow Boon Kuan Mr Chung Kwong Yuew Ms Heidi Goh Prof James Goh Dr Lai Choon Sheen Dr Christopher Lam Assoc Prof Leo Hwa Liang Dr Lin Jianhua Dr Leonard Loh Assoc Prof Eddie Ng Yin Kwee Dr Ong Siew Hwa Dr Padmanabhan Saravanan Mr Peh Ruey Feng Ms Celine Tan Prof Tan Puay Hoon Ms Wang Dan Dr Sidney Yee Dr Zhou Zhihong	<i>Medtronic International Ltd</i> <i>Temasek Polytechnic (BioMedical Engineering Faculty)</i> <i>Singapore Manufacturing Federation (Medical Technology Industry Group)</i> <i>Biomedical Engineering Society (Singapore)</i> <i>Eu Yan Sang International Ltd</i> <i>Health Sciences Authority</i> <i>National University of Singapore</i> <i>TüV SüD PSB Pte Ltd</i> <i>Nanyang Polytechnic</i> <i>Nanyang Technological University</i> <i>Acumen Research Laboratories Pte Ltd</i> <i>Temasek Polytechnic (Centre of Innovation for Complementary Health Products)</i> <i>Advent Access Pte Ltd</i> <i>Enterprise Singapore</i> <i>Singapore Health Services Pte Ltd</i> <i>Biosensors International Group</i> <i>Diagnostics Development (DxD) Hub</i> <i>Singapore Bioimaging Consortium</i>

## SS IEC 60601-1-2 : 2018

The Technical Committee on Medical Devices, appointed by the Biomedical and Health Standards Committee, consists of representatives from the following organisations:

	<b>Name</b>	<b>Capacity</b>
<b>Chairman</b>	: Prof James Goh	<i>Individual Capacity</i>
<b>Secretary</b>	: Ms Iris Peng	<i>Singapore Manufacturing Federation – Standards Development Organisation</i>
<b>Members</b>	: Ms Chua Chui Khim	<i>Becton Dickinson Medical (S) Pte Ltd</i>
	Dr Christopher Lam	<i>Health Sciences Authority</i>
	Assoc Prof Leo Hwa Liang	<i>Biomedical Engineering Society (Singapore)</i>
	Dr Lim Jing	<i>Osteopore International Pte Ltd</i>
	Ms Iris Tan	<i>Advent Access Pte Ltd</i>

The National Mirror Working Group on IEC SC62D MT20, appointed by the Technical Committee to assist in the preparation of this standard, comprises the following experts who contribute in their individual capacity:

	<b>Name</b>
<b>Convenor</b>	: Dr Christian Gert Bluechel
<b>Secretary</b>	: Mr Kevin Tan
<b>Members</b>	: Mr Peter Haywood
	Dr Ho Teck Tuak
	Mr Watson Ong
	Mr Peh Ruey Feng
	Prof Tong Yen Wah

The organisations in which the experts of the National Mirror Working Group are involved are:

*Advent Access Pte Ltd*  
*AWAK Technologies Pte Ltd*  
*Kidney Dialysis Foundation*  
*National University of Singapore*  
*Temasek Polytechnic*  
*The Institution of Engineers, Singapore*

(blank page)

## CONTENTS

NATIONAL FOREWORD .....	9
FOREWORD .....	10
INTRODUCTION .....	13
1 Scope, object and related standards .....	15
1.1 * Scope .....	15
1.2 Object.....	15
1.3 Related standards .....	15
1.3.1 IEC 60601-1 .....	15
1.3.2 Particular standards .....	15
2 Normative references .....	16
3 Terms and definitions .....	18
4 General requirements.....	21
4.1 RISK MANAGEMENT PROCESS for ME EQUIPMENT and ME SYSTEMS .....	21
4.2 * Non-ME EQUIPMENT used in an ME SYSTEM .....	21
4.3 General test conditions .....	22
4.3.1 * Configurations .....	22
4.3.2 Artificial hand .....	22
4.3.3 * Power input voltages and frequencies .....	23
5 ME EQUIPMENT and ME SYSTEMS identification, marking and documents .....	25
5.1 Additional requirements for marking on the outside of ME EQUIPMENT and ME SYSTEMS that are specified for use only in a shielded location SPECIAL ENVIRONMENT .....	25
5.2 ACCOMPANYING DOCUMENTS .....	25
5.2.1 Instructions for use.....	25
5.2.2 Technical description .....	27
6 Documentation of the tests.....	28
6.1 General .....	28
6.2 Test plan .....	28
6.3 Test report.....	28
7 ELECTROMAGNETIC EMISSIONS requirements for ME EQUIPMENT and ME SYSTEMS.....	29
7.1 Protection of radio services and other equipment .....	29
7.1.1 * General.....	29
7.1.2 Operating modes.....	29
7.1.3 Multimedia equipment .....	29
7.1.4 * Subsystems .....	29
7.1.5 ME EQUIPMENT and ME SYSTEMS specified for use only in a shielded location SPECIAL ENVIRONMENT .....	29
7.1.6 ME EQUIPMENT and ME SYSTEMS that include radio equipment .....	30
7.1.7 * ME EQUIPMENT whose main functions are performed by motors and switching or regulating devices .....	30
7.1.8 ME EQUIPMENT and ME SYSTEMS containing X-ray generators.....	30
7.1.9 PATIENT physiological simulation.....	30
7.1.10 Artificial hand .....	31
7.1.11 PATIENT-coupled cables .....	31

7.1.12	PERMANENTLY INSTALLED LARGE ME EQUIPMENT and LARGE ME SYSTEMS .....	31
7.2	Protection of the PUBLIC MAINS NETWORK.....	31
7.2.1	* Harmonic distortion .....	31
7.2.2	* Voltage fluctuations and flicker .....	31
7.3	EMISSIONS requirements summary .....	32
8	Electromagnetic IMMUNITY requirements for ME EQUIPMENT and ME SYSTEMS .....	32
8.1	* General .....	32
8.2	PATIENT physiological simulation.....	35
8.3	Termination of PATIENT-COUPLED parts .....	36
8.4	HAND-HELD ME EQUIPMENT and parts intended to be HAND-HELD .....	36
8.5	* Subsystems .....	36
8.6	PERMANENTLY INSTALLED LARGE ME EQUIPMENT and LARGE ME SYSTEMS.....	36
8.7	* Operating modes.....	37
8.8	* Non-ME EQUIPMENT .....	37
8.9	* IMMUNITY TEST LEVELS .....	37
8.10	* IMMUNITY to proximity fields from RF wireless communications equipment .....	45
9	* Test report.....	47
Annex A	(informative) General guidance and rationale .....	50
A.1	Safety and performance.....	50
A.2	Testing of normally non-observable functions .....	50
A.3	Rationale for particular clauses and subclauses .....	50
Annex B	(informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS .....	64
B.1	Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts.....	64
B.2	ACCOMPANYING DOCUMENTS, instructions for use .....	64
B.3	ACCOMPANYING DOCUMENTS, technical description .....	64
Annex C	(informative) Guidance in classification according to CISPR 11 .....	66
C.1	General .....	66
C.2	Separation into groups .....	66
C.3	Division into classes .....	67
Annex D	(informative) Guidance in the application of IEC 60601-1-2 to particular standards .....	68
D.1	General .....	68
D.2	Recommended modifications .....	68
D.2.1	Testing requirements.....	68
D.2.2	ACCOMPANYING DOCUMENTS.....	68
D.3	Cautions .....	68
Annex E	(informative) Determination of IMMUNITY TEST LEVELS for SPECIAL ENVIRONMENTS .....	70
E.1	General .....	70
E.2	Summary of method for E.1 a) .....	73
E.3	Summary of method for E.1 b), c) and d) .....	73
E.4	Determination of EM DISTURBANCE level reduction .....	73
E.5	Assessment of EM DISTURBANCE sources .....	73

E.6	Reasonably foreseeable maximum EM DISTURBANCE levels.....	74
E.7	Determination of IMMUNITY TEST LEVELS .....	74
E.8	RF radiators in SPECIAL ENVIRONMENTS .....	75
E.9	Examples of mitigations and special conditions .....	75
Annex F (informative) RISK MANAGEMENT for BASIC SAFETY and ESSENTIAL PERFORMANCE with regard to ELECTROMAGNETIC DISTURBANCES .....		77
F.1	General .....	77
F.2	General requirements for RISK MANAGEMENT .....	78
F.3	RISK ANALYSIS .....	79
F.4	RISK EVALUATION .....	83
F.5	RISK CONTROL .....	83
F.5.1	RISK CONTROL option analysis .....	83
F.5.2	Implementation of RISK CONTROL measure(s) .....	83
F.5.3	RESIDUAL RISK EVALUATION .....	84
F.5.4	RISK/benefit analysis .....	84
F.5.5	RISKS arising from RISK CONTROL measures .....	84
F.5.6	Completeness of RISK CONTROL.....	84
F.6	Evaluation of overall RESIDUAL RISK acceptability .....	85
F.7	RISK MANAGEMENT report.....	85
F.8	Production and post-production information .....	85
Annex G (informative) Guidance: Test plan .....		86
G.1	Test plan contents .....	86
Annex H (informative) PATIENT-coupled cables EMISSIONS.....		88
H.1	* Protection of other equipment from PATIENT cable conducted EMISSIONS .....	88
H.2	Test method .....	88
H.3	Rationale .....	88
Annex I (informative) Identification of IMMUNITY pass/fail criteria .....		90
I.1	General .....	90
I.2	IMMUNITY pass/fail criteria principles .....	90
I.2.1	General .....	90
I.2.2	IMMUNITY pass/fail criteria for non-ME EQUIPMENT used in an ME SYSTEM.....	90
I.2.3	IMMUNITY pass/fail criteria determination.....	90
I.3	IMMUNITY pass/fail criteria examples .....	91
I.3.1	General examples .....	91
I.3.2	Example of IMMUNITY pass/fail criteria for a radiological table system .....	92
Bibliography .....		94
Index of defined terms used in this collateral standard.....		97
Figure 1 – RC element of the artificial hand.....		23
Figure 2 – PORTS of ME EQUIPMENT and ME SYSTEMS.....		32
Figure 3 – Examples of environments of INTENDED USE .....		39
Figure A.1 – Examples of PORTS (from IEC 61000-6-1:2005).....		55



Figure A.2 – IEC 61000-4-2 Figure A.1 – Maximum values of electrostatic voltages to which OPERATORS can be charged while in contact with the materials mentioned in A.2 ..... 61

Figure E.1 – Test plan development flow when SPECIAL ENVIRONMENTS are known ..... 71

Figure E.2 – Sub-process for determination of IMMUNITY TEST LEVELS for SPECIAL ENVIRONMENTS ..... 72

Figure F.1 – Function of this collateral standard in the RISK MANAGEMENT PROCESS ..... 77

Figure F.2 – Examples of multiple VERIFICATION methods for improving confidence in RISK levels ..... 78

Figure H.1 – Setup for PATIENT-COUPLED cables conducted EMISSIONS test for ME EQUIPMENT and ME SYSTEMS that conform to IEC 60601-2-27 ..... 89

  

Table 1 – Power input voltages and frequencies during the tests (1 of 2) ..... 24

Table 2 – EMISSION limits per environment ..... 32

Table 3 – Procedure for continuing to test ME EQUIPMENT or ME SYSTEMS that are damaged by an IMMUNITY test signal ..... 33

Table 4 – \* ENCLOSURE PORT ..... 40

Table 5 – \* Input a.c. power PORT (1 of 2) ..... 41

Table 6 – Input d.c. power PORT ..... 43

Table 7 – \* PATIENT coupling PORT ..... 44

Table 8 – Signal input/output parts PORT ..... 45

Table 9 – Test specifications for ENCLOSURE PORT IMMUNITY to RF wireless communications equipment ..... 46

Table 10 – \* Minimum test report contents (1 of 2) ..... 48

Table A.1 – IEC/TR 61000-2-5 information considered in specifying IMMUNITY TEST LEVELS for each IMMUNITY TEST ..... 57

Table B.1 – Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts ..... 64

Table B.2 – ACCOMPANYING DOCUMENTS, instructions for use ..... 64

Table B.3 – ACCOMPANYING DOCUMENTS, technical description ..... 65

Table E.1 – Examples of specific mitigations / environmental conditions ..... 76

Table F.1 – Examples of EM phenomena that should be considered in a RISK ANALYSIS ..... 80

Table G.1 – Recommended minimum test plan contents (1 of 2) ..... 86

Table H.1 – PATIENT-COUPLED conducted EMISSIONS recommended limit ..... 88

Table I.1 – Example of IMMUNITY pass criteria for a radiological table system ..... 93

## National Foreword

This Singapore Standard was prepared by the National Mirror Working Group on IEC SC62D MT20 appointed by the Technical Committee on Medical Devices under the direction of the Biomedical and Health Standards Committee.

This standard is identical with IEC 60601-1-2:2014, “Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic disturbances – Requirements and tests”, published by the International Electrotechnical Commission.

Attention is drawn to the following:

1. Where appropriate, the words “International Standard” shall be read as “Singapore Standard”.
2. Where reference to a particular part of IEC 60601 is made, the appropriate Singapore Standard (which is an identical adoption of that part of IEC 60601) shall be referenced. The reference to “ISO 14971:2007” shall be replaced by “SS ISO 14971:2017”.
3. The comma has been used throughout as a decimal marker whereas in Singapore Standards it is a practice to use a full point on the baseline as the decimal marker.

Attention is drawn to the possibility that some of the elements of this Singapore Standard may be the subject of patent rights. Enterprise Singapore shall not be held responsible for identifying any or all of such patent rights.

### NOTE

1. *Singapore Standards (SSs) and Technical References (TRs) are reviewed periodically to keep abreast of technical changes, technological developments and industry practices. The changes are documented through the issue of either amendments or revisions.*
2. *An SS or TR is voluntary in nature except when it is made mandatory by a regulatory authority. It can also be cited in contracts making its application a business necessity. Users are advised to assess and determine whether the SS or TR is suitable for their intended use or purpose. If required, they should refer to the relevant professionals or experts for advice on the use of the document. Enterprise Singapore shall not be liable for any damages whether directly or indirectly suffered by anyone or any organisation as a result of the use of any SS or TR.*
3. *Compliance with a SS or TR does not exempt users from any legal obligations.*

**INTERNATIONAL ELECTROTECHNICAL COMMISSION**

**MEDICAL ELECTRICAL EQUIPMENT –**

**Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests**

**FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-1-2 has been prepared by IEC subcommittee 62A: Common aspects of electrical equipment used in medical practice of IEC technical committee 62: Electrical equipment in medical practice.

This fourth edition cancels and replaces the third edition of IEC 60601-1-2, published in 2007, and constitutes a technical revision.

This fourth edition constitutes a collateral standard to IEC 60601-1: *Medical electrical equipment – Part 1: General requirements for safety and essential performance* hereafter referred to as the general standard.

The most significant changes with respect to the previous edition include the following modifications:

- specification of IMMUNITY TEST LEVELS according to the environments of INTENDED USE, categorized according to locations that are harmonized with IEC 60601-1-11: the professional healthcare facility environment, the HOME HEALTHCARE ENVIRONMENT and SPECIAL ENVIRONMENTS;
- specification of tests and test levels to improve the safety of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS when PORTABLE RF communications equipment is used closer to the MEDICAL ELECTRICAL EQUIPMENT than was recommended based on the IMMUNITY TEST LEVELS that were specified in the third edition;
- specification of IMMUNITY tests and IMMUNITY TEST LEVELS according to the PORTS of the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM;
- specification of IMMUNITY TEST LEVELS based on the reasonably foreseeable maximum level of ELECTROMAGNETIC DISTURBANCES in the environments of INTENDED USE, resulting in some IMMUNITY TEST LEVELS that are higher than in the previous edition; and
- better harmonization with the RISK concepts of BASIC SAFETY and ESSENTIAL PERFORMANCE, including deletion of the defined term “life-supporting”;

and the following additions:

- guidance for determination of IMMUNITY TEST LEVELS for SPECIAL ENVIRONMENTS;
- guidance for adjustment of IMMUNITY TEST LEVELS when special considerations of mitigations or INTENDED USE are applicable;
- guidance on RISK MANAGEMENT for BASIC SAFETY and ESSENTIAL PERFORMANCE with regard to ELECTROMAGNETIC DISTURBANCES; and
- guidance on identification of IMMUNITY pass/fail criteria.

The text of this collateral standard is based on the following documents:

FDIS	Report on voting
62A/916/FDIS	62A/924/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. radiological equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. ALARM SYSTEMS).

In this collateral standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this collateral standard, the term

- “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 1 includes 1.1, 1.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 1.1, 1.2 and 1.3.1 are all subclauses of Clause 1).

References to clauses within this collateral standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this collateral standard are by number only.

In this collateral standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this collateral standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this collateral standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this collateral standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this collateral standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex A.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip them for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for mandatory implementation nationally not earlier than 3 years from the date of publication.

**IMPORTANT – The “colour inside” logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.**

## INTRODUCTION

The need for establishing specific standards for BASIC SAFETY and ESSENTIAL PERFORMANCE with regard to ELECTROMAGNETIC DISTURBANCES for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS is well recognized.

The requirements and tests specified by this collateral standard are generally applicable to MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS as defined in 3.63 and 3.64 in the general standard. For certain types of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, these requirements might need to be modified by the special requirements of a particular standard. Writers of particular standards are encouraged to refer to Annex D for guidance in the application of this collateral standard.

MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS are expected to provide their BASIC SAFETY and ESSENTIAL PERFORMANCE without interfering with other equipment and systems in the ELECTROMAGNETIC ENVIRONMENTS in which they are intended by their MANUFACTURER to be used. The application of ELECTROMAGNETIC EMISSION standards is essential for the protection of:

- safety services;
- other MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS;
- non-ME EQUIPMENT (e.g. computers);
- telecommunications (e.g. radio/TV, telephone, radio-navigation).

Of even more importance, the application of ELECTROMAGNETIC IMMUNITY standards is essential to ensure safety of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS. To ensure safety, MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS are expected to provide their BASIC SAFETY and ESSENTIAL PERFORMANCE in the ELECTROMAGNETIC ENVIRONMENTS of INTENDED USE throughout their EXPECTED SERVICE LIFE.

This collateral standard specifies IMMUNITY TEST LEVELS for safety for ME EQUIPMENT and ME SYSTEMS intended by their MANUFACTURER for use in the professional healthcare facility environment or the HOME HEALTHCARE ENVIRONMENT. It recognizes that RF wireless communications equipment can no longer be prohibited from most PATIENT ENVIRONMENTS because in many cases it has become essential to the efficient provision of healthcare. This collateral standard also recognizes that, for certain SPECIAL ENVIRONMENTS, higher or lower IMMUNITY TEST LEVELS than those specified for the professional healthcare facility environment and the HOME HEALTHCARE ENVIRONMENT might be appropriate. This collateral standard provides guidance in determining appropriate IMMUNITY TEST LEVELS for SPECIAL ENVIRONMENTS.

The IMMUNITY TEST LEVELS specified for BASIC SAFETY and ESSENTIAL PERFORMANCE are based on the reasonably foreseeable maximum of the ELECTROMAGNETIC DISTURBANCE phenomena in the applicable environments of INTENDED USE.

Not all ELECTROMAGNETIC DISTURBANCE phenomena are covered by this collateral standard, as it is not practical to do so. MANUFACTURERS of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS need to address this during their RISK ASSESSMENT and evaluate if other ELECTROMAGNETIC DISTURBANCE phenomena could make their product unsafe. This evaluation should be based on the environments of INTENDED USE and the reasonably foreseeable maximum levels of ELECTROMAGNETIC DISTURBANCES expected throughout the EXPECTED SERVICE LIFE.

This collateral standard recognizes that the MANUFACTURER has the responsibility to design and perform VERIFICATION of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS to meet the requirements of this collateral standard and to disclose information to the RESPONSIBLE ORGANIZATION or OPERATOR so that the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM will remain safe throughout its EXPECTED SERVICE LIFE.

This collateral standard provides guidance in incorporating considerations regarding ELECTROMAGNETIC DISTURBANCES into the RISK MANAGEMENT PROCESS.

This collateral standard is based on existing IEC standards prepared by subcommittee 62A, technical committee 77 (ELECTROMAGNETIC COMPATIBILITY between electrical equipment including networks), ISO (International standards organization), and CISPR (International special committee on radio interference).

**MEDICAL ELECTRICAL EQUIPMENT –**

**Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests**

**1 Scope, object and related standards**

**1.1 \* Scope**

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, hereafter referred to as ME EQUIPMENT and ME SYSTEMS.

This collateral standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of ME EQUIPMENT and ME SYSTEMS in the presence of ELECTROMAGNETIC DISTURBANCES and to ELECTROMAGNETIC DISTURBANCES emitted by ME EQUIPMENT and ME SYSTEMS.

BASIC SAFETY with regard to ELECTROMAGNETIC DISTURBANCES is applicable to all ME EQUIPMENT and ME SYSTEMS.