

SS IEC 60601-1-3 : 2018 IEC 60601-1-3:2008+A1:2013, IDT

(ICS 11.040.50; 13.280)

# SINGAPORE STANDARD

# Medical electrical equipment

 Part 1-3: General requirements for basic safety and essential performance – Collateral Standard: Radiation protection in diagnostic X-ray equipment



Published by



**SS IEC 60601-1-3 : 2018** IEC 60601-1-3:2008+A1:2013, IDT (ICS 11.040.50; 13.280)

#### SINGAPORE STANDARD

# Medical electrical equipment

Part 1-3: General requirements for basic safety and essential performance –
 Collateral Standard: Radiation protection in diagnostic X-ray equipment

All rights reserved. Unless otherwise specified, no part of this Singapore Standard may be reproduced or utilised in any form or by any means, electronic or mechanical, including photocopying and microfilming, without permission in writing from Enterprise Singapore. Request for permission can be sent to: standards@enterprisesg.gov.sg.

© IEC 2008 – All rights reserved © Enterprise Singapore 2018

ISBN 978-981-48-3528-2

This Singapore Standard was approved on 19 October 2018 by the Biomedical and Health Standards Committee under the purview of the Singapore Standards Council.

First published, 2019

The Biomedical and Health Standards Committee, appointed by the Standards Council, consists of the following members:

		Name	Capacity
Chairman	:	Dr Yong Chern Chet	Individual Capacity
Deputy Chairmen	:	Mr Vincent Cheung Ms Selina Seah Ms Wong Woei Jiuang	Individual Capacity Changi General Hospital Health Sciences Authority
Advisor	:	Ms Jacqueline Monteiro	Individual Capacity
Secretary	:	Ms Iris Peng	Singapore Manufacturing Federation – Standards Development Organisation
Members	:	Mr Alec Chow Boon Kuan	Medtronic International Ltd
		Mr Chung Kwong Yuew	Temasek Polytechnic (BioMedical Engineering Faculty)
		Ms Heidi Goh	Singapore Manufacturing Federation (Medical Technology Industry Group)
		Prof James Goh	Biomedical Engineering Society (Singapore)
		Dr Lai Choon Sheen	Eu Yan Sang International Ltd
		Dr Christopher Lam	Health Sciences Authority
		Assoc Prof Leo Hwa Liang	National University of Singapore
		Dr Lin Jianhua	TüV SüD PSB Pte Ltd
		Dr Leonard Loh	Nanyang Polytechnic
		Assoc Prof Eddie Ng Yin Kwee	Nanyang Technological University
		Dr Ong Siew Hwa	Acumen Research Laboratories Pte Ltd
		Dr Padmanabhan Saravanan	Temasek Polytechnic (Centre of Innovation for Complementary Health Products)
		Mr Peh Ruey Feng	Advent Access Pte Ltd
		Ms Celine Tan	Enterprise Singapore
		Prof Tan Puay Hoon	Singapore Health Services Pte Ltd
		Ms Wang Dan	Biosensors International Group
		Dr Sidney Yee	Diagnostics Development (DxD) Hub
		Dr Zhou Zhihong	Singapore Bioimaging Consortium

The Technical Committee on Medical Devices, appointed by the Biomedical and Health Standards Committee, consists of representatives from the following organisations:

		Name	Capacity
Chairman	:	Prof James Goh	Individual Capacity
Secretary	:	Ms Iris Peng	Singapore Manufacturing Federation – Standards Development Organisation
Members	:	Ms Chua Chui Khim Dr Christopher Lam	Becton Dickinson Medical (S) Pte Ltd Health Sciences Authority
		Assoc Prof Leo Hwa Liang	Biomedical Engineering Society (Singapore)
		Dr Lim Jing	Osteopore International Pte Ltd
		Ms Iris Tan	Advent Access Pte Ltd

The National Mirror Working Group on IEC SC62D MT20, appointed by the Technical Committee to assist in the preparation of this standard, comprises the following experts who contribute in their individual capacity:

#### Name

Convenor : Dr Christian Gert Bluechel

Secretary Mr Kevin Tan

Members : Mr Peter Haywood

> Dr Ho Teck Tuak Mr Watson Ong Mr Peh Ruey Feng Prof Tong Yen Wah

The organisations in which the experts of the National Mirror Working Group are involved are:

Advent Access Pte Ltd AWAK Technologies Pte Ltd Kidney Dialysis Foundation National University of Singapore Temasek Polytechnic

The Institution of Engineers, Singapore

(blank page)

# **CONTENTS**

NA	TIONA	AL FOR	EWORD	8
FO	REWO	)RD		9
INT	RODU	JCTION	<b>1</b>	12
INT	RODU	JCTION	TO AMENDMENT 1	14
1	Scope, object and related standards			15
	1.1 Scope			15
	1.2	Object		15
	1.3	Related standards		15
		1.3.1	IEC 60601-1	15
		1.3.2	Particular standards	16
2	Norm	ative re	eferences	16
3	Term	s and d	lefinitions	16
4	Gene	ral requ	uirements	27
	4.1	Staten	nent of compliance	27
	4.2	Compo	osition of reference materials	27
5	ME E	QUIPMEI	NT identification, marking and documents	27
	5.1	Markir	ng on the outside of ME EQUIPMENT or ME EQUIPMENT parts	27
		5.1.1		
		5.1.2	Marking requirements in subclauses	27
	5.2	Accon	MPANYING DOCUMENTS	27
		5.2.1	References in subclauses	28
		5.2.2	Dosimetric calibration	28
		5.2.3	General requirements for the reference of subassemblies and ACCESSORIES	29
		5.2.4	Instructions for use	29
6	RADIA	ATION M	nanagement	31
	6.1	Gener	al	31
	6.2	Initiati	on and termination of the IRRADIATION	31
		6.2.1	Normal initiation and termination of the IRRADIATION	31
		6.2.2	Safety measures against failure of normal termination of the IRRADIATION	31
	6.3	RADIAT	TION dose and RADIATION QUALITY	31
		6.3.1	Adjustment of RADIATION dose and RADIATION QUALITY	31
		6.3.2	Reproducibility of the RADIATION output	32
	6.4	Indication of operational states		
		6.4.1	Indication of the X-RAY SOURCE ASSEMBLY selected	32
		6.4.2	Indication of LOADING STATE	
		6.4.3	Indication of LOADING FACTORS and MODES OF OPERATION	
		6.4.4	Indication of automatic modes	
	_	6.4.5	Dosimetric indications	
	6.5		IATIC CONTROL SYSTEM	
	6.6	SCATTERED RADIATION reduction		34

	6.7	Imaging performance	34	
		6.7.1 General	34	
		6.7.2 System performance	34	
		6.7.3 Nominal focal spot value		
		6.7.4 RADIATION DETECTOR OF X-RAY IMAGE RECEPTOR	34	
7	RADI	ATION QUALITY	35	
	7.1	HALF-VALUE LAYERS and TOTAL FILTRATION IN X-RAY EQUIPMENT	35	
	7.2	Waveform of the X-RAY TUBE VOLTAGE	36	
	7.3	Indication of FILTER properties	36	
	7.4	Test for FILTRATION by irremovable materials	37	
	7.5	Test for ADDED FILTERS and materials	37	
	7.6	Test for HALF-VALUE LAYER	37	
8	Limitation of the extent of the X-RAY BEAM and relationship between X-RAY FIELD and IMAGE RECEPTION AREA			
	8.1	General		
	8.2	Enclosure of X-RAY TUBES		
	8.3	Limiting DIAPHRAGM in X-RAY TUBE ASSEMBLIES		
	8.4	Confinement of EXTRA-FOCAL RADIATION		
	8.5	Relationship between X-RAY FIELD and IMAGE RECEPTION AREA		
	0.0	8.5.1 General		
		8.5.2 * FOCAL SPOT TO IMAGE RECEPTOR DISTANCE		
		8.5.3 Correspondence between X-RAY FIELD and EFFECTIVE IMAGE RECEPTION AREA		
		8.5.4 Positioning of the PATIENT and restriction of the irradiated area		
9	FOCA	AL SPOT TO SKIN DISTANCE		
	9.1	General		
	9.2	Information in the ACCOMPANYING DOCUMENTS		
10	•	NUATION of the X-RAY BEAM between the PATIENT and the X-RAY IMAGE	00	
		PTOR	39	
	10.1	General	39	
		Information in the ACCOMPANYING DOCUMENTS		
11		ection against RESIDUAL RADIATION		
		otection against LEAKAGE RADIATION		
	12.1	General	40	
	12.2	Mounting of X-RAY SOURCE ASSEMBLIES and X-RAY IMAGING ARRANGEMENTS	41	
	12.3	Statement of reference LOADING conditions	41	
	12.4	LEAKAGE RADIATION in the LOADING STATE	41	
	12.5 LEAKAGE RADIATION when not in the LOADING STATE			
13		ection against STRAY RADIATION		
		General		
		Control of X-ray equipment from a protected area		
		Protection by distance		
		* Designated SIGNIFICANT ZONES OF OCCUPANCY		
		Handgrips and control devices		
		* Test for STRAY RADIATION	11	

Annex A (informative) General guidance and rationale	46
Annex B (normative) Values of the series R'10 and R'20, ISO 497	48
Annex C (informative) Mapping between this Edition 2 of IEC 60601-1-3 and Edition 1	49
Bibliography	51
Index of defined terms used in this collateral standard	53
Figure 1 – Example of presentation of data on STRAY RADIATION	45
Table 1 – Subclauses containing requirements for marking	27
Table 2 – Subclauses requiring statements in ACCOMPANYING DOCUMENTS	28
Table 3 _ HALE-VALUE LAVERS IN Y-DAY COULDMENT	35

#### **National Foreword**

This Singapore Standard was prepared by the National Mirror Working Group on IEC SC62D MT20 appointed by the Technical Committee on Medical Devices under the direction of the Biomedical and Health Standards Committee.

This standard is identical with IEC 60601-1-3:2008+A1:2013, "Medical electrical equipment – Part 1-3: General requirements for basic safety and essential performance – Collateral Standard: Radiation protection in diagnostic X-ray equipment", published by the International Electrotechnical Commission.

Attention is drawn to the following:

- 1. Where appropriate, the words "International Standard" shall be read as "Singapore Standard".
- 2. Where reference to a particular part of IEC 60601 is made, the appropriate Singapore Standard (which is an identical adoption of that part of IEC 60601) shall be referenced.
- 3. The comma has been used throughout as a decimal marker whereas in Singapore Standards it is a practice to use a full point on the baseline as the decimal marker.

Attention is drawn to the possibility that some of the elements of this Singapore Standard may be the subject of patent rights. Enterprise Singapore shall not be held responsible for identifying any or all of such patent rights.

#### NOTE

- 1. Singapore Standards (SSs) and Technical References (TRs) are reviewed periodically to keep abreast of technical changes, technological developments and industry practices. The changes are documented through the issue of either amendments or revisions.
- 2. An SS or TR is voluntary in nature except when it is made mandatory by a regulatory authority. It can also be cited in contracts making its application a business necessity. Users are advised to assess and determine whether the SS or TR is suitable for their intended use or purpose. If required, they should refer to the relevant professionals or experts for advice on the use of the document. Enterprise Singapore shall not be liable for any damages whether directly or indirectly suffered by anyone or any organisation as a result of the use of any SS or TR.
- 3. Compliance with a SS or TR does not exempt users from any legal obligations.

#### INTERNATIONAL ELECTROTECHNICAL COMMISSION

#### MEDICAL ELECTRICAL EQUIPMENT -

Part 1-3: General requirements for basic safety and essential performance – Collateral Standard: Radiation protection in diagnostic X-ray equipment

#### **FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of IEC 60601-1-3 consists of the second edition (2008) [documents 62B/673/FDIS and 62B/683/RVD] and its amendment 1 (2013) [documents 62B/895/CDV and 62B/907/RVC]. It bears the edition number 2.1.

The technical content is therefore identical to the base edition and its amendment and has been prepared for user convenience. A vertical line in the margin shows where the base publication has been modified by amendment 1. Additions and deletions are displayed in red, with deletions being struck through.

International standard IEC 60601-1-3 has been prepared by IEC subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This edition constitutes a collateral standard to IEC 60601-1: *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance* hereafter referred to as the general standard.

This edition has been restructured and aligned to IEC 60601-1(2005) and focussed on general requirements for RADIATION PROTECTION that apply to all diagnostic X-RAY EQUIPMENT. Requirements particular to specific equipment have been removed and will be covered in particular standards. For a description of the changes, see the mapping in Annex C.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. RADIOLOGICAL equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. alarm systems).

In this collateral standard, the following print types are used:

- requirements and definitions: roman type.
- test specifications: italic type.
- informative material appearing outside of tables, such as notes, examples and references: in smaller type.
   Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- "clause" means one of the thirteen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term "Clause" followed by the clause number. References to subclauses within this collateral standard are by number only.

In this standard, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this standard;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;

 - "may" is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (\*).

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendment will remain unchanged until the maintenance result date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- · reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended

IMPORTANT – The "colour inside" logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.

#### INTRODUCTION

The requirements in this collateral Standard concern protective measures to be taken by the MANUFACTURER in the design and construction of medical diagnostic X-RAY EQUIPMENT and its subassemblies. They relate to the application of the X-RADIATION generated, both deliberately and incidentally, in fulfilling the medical purpose of the EQUIPMENT. Additional measures are necessary to regulate the generation processes themselves. These are described in the general requirements for safety, IEC 60601-1, and, where appropriate, in particular requirements for the EQUIPMENT concerned. The second edition of this collateral standard is focused on general requirements for RADIATION PROTECTION. The aim of the revision was to restrict to those requirements that apply to all diagnostic X-RAY EQUIPMENT. In consequence, most of the clauses have been reduced compared with the first edition of this standard, owing to the exclusion of content specific to projection RADIOGRAPHY and RADIOSCOPY. Implementation shall be considered in the RISK MANAGEMENT process or by using particular standards.

The recommended principles governing the use of RADIATION for medical purposes, as stated in Publication 60 of the International Commission on Radiological Protection (ICRP)[17]<sup>1)</sup>, Chapter 4, have been taken into account. The implementation of these principles is essentially determined in the prevailing circumstances at the point of use. It requires judgements to be made by the user and the establishment of measures and working practices part of which are connected with the construction of EQUIPMENT. The requirements in this collateral Standard are intended to be consistent with generally accepted good practice in the administration of X-RADIATION in medicine.

In some cases, the formulation of the requirements is deliberately designed to provide scope for accommodating local laws and regulations at the time of installation and commissioning. Several of the requirements include provisions for relevant technical information to be included in ACCOMPANYING DOCUMENTS.

RESPONSIBLE ORGANIZATIONS for medical diagnostic X-RAY EQUIPMENT should be aware that effective protection against IONIZING RADIATION requires the consideration of many aspects additional to the construction of the EQUIPMENT. Among these are the following:

- compatibility of components and correct installation of EQUIPMENT;
- the protective properties of rooms where X-RAY EQUIPMENT is installed;
- measures for monitoring and maintaining the safety and effectiveness of EQUIPMENT throughout its life, with particular attention to components that can deteriorate progressively with time and use;
- the need in appropriate circumstances for PROTECTIVE CLOTHING to be worn by staff and for suitable devices to be used to protect PATIENTS;
- the keeping of appropriate records concerning the usage of the EQUIPMENT and the results
  of tests, with systematic review and the application of corrective action when necessary;
- the training of staff in the principles of RADIATION PROTECTION and in the correct use of EQUIPMENT, including any PROTECTIVE DEVICES provided.

Further advice on these aspects can be found in ICRP Publications 33[15], 34[16], 60[17], 73[18], 85[21], 87[22] and 93[23].

<sup>1)</sup> Figures in square brackets refer to the Bibliography.

Readers of this collateral standard are reminded that, in accordance with IEC 60601-1, Clause 5, all the test procedures described are TYPE TESTS, intended to be carried out in a dedicated testing environment in order to determine compliance. Tests to be carried out by MANUFACTURERS to ensure compliance during production or installation and tests for detecting non-compliance subsequently to delivery, are not included.

# **INTRODUCTION TO AMENDMENT 1**

The purpose of the first amendment to IEC 60601-1-3:2008 is to introduce changes to reference the first amendment (2012) to IEC 60601-1:2005.

### **MEDICAL ELECTRICAL EQUIPMENT -**

Part 1-3: General requirements for basic safety and essential performance – Collateral Standard:

Radiation protection in diagnostic X-ray equipment

### 1 Scope, object and related standards

#### 1.1 Scope

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, hereafter referred to as ME EQUIPMENT and ME SYSTEMS.

This collateral standard applies to X-RAY EQUIPMENT and to subassemblies of such equipment, where RADIOLOGICAL IMAGES of a human PATIENT are used for diagnosis, planning or guidance of medical procedures.