

SINGAPORE STANDARD**Medical electrical equipment**

– Part 1-8: General requirements for basic safety and essential performance – Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems

Published by

**Enterprise
Singapore**

SS IEC 60601-1-8 : 2018

IEC 60601-1-8:2006+A1:2012, IDT
(ICS 11.040.01)

SINGAPORE STANDARD

Medical electrical equipment

– Part 1-8: General requirements for basic safety and essential performance –
Collateral Standard: General requirements, tests and guidance for alarm
systems in medical electrical equipment and medical electrical systems

All rights reserved. Unless otherwise specified, no part of this Singapore Standard may be reproduced or utilised in any form or by any means, electronic or mechanical, including photocopying and microfilming, without permission in writing from Enterprise Singapore. Request for permission can be sent to: standards@enterprisesg.gov.sg.

© IEC 2006 – All rights reserved
© Enterprise Singapore 2018

ISBN 978-981-48-3530-5

This Singapore Standard was approved on 19 October 2018 by the Biomedical and Health Standards Committee under the purview of the Singapore Standards Council.

First published, 2019

The Biomedical and Health Standards Committee, appointed by the Standards Council, consists of the following members:

	Name	Capacity
Chairman	: Dr Yong Chern Chet	<i>Individual Capacity</i>
Deputy Chairmen	: Mr Vincent Cheung Ms Selina Seah Ms Wong Woei Jiuang	<i>Individual Capacity</i> <i>Changi General Hospital</i> <i>Health Sciences Authority</i>
Advisor	: Ms Jacqueline Monteiro	<i>Individual Capacity</i>
Secretary	: Ms Iris Peng	<i>Singapore Manufacturing Federation – Standards Development Organisation</i>
Members	: Mr Alec Chow Boon Kuan Mr Chung Kwong Yuew Ms Heidi Goh Prof James Goh Dr Lai Choon Sheen Dr Christopher Lam Assoc Prof Leo Hwa Liang Dr Lin Jianhua Dr Leonard Loh Assoc Prof Eddie Ng Yin Kwee Dr Ong Siew Hwa Dr Padmanabhan Saravanan Mr Peh Ruey Feng Ms Celine Tan Prof Tan Puay Hoon Ms Wang Dan Dr Sidney Yee Dr Zhou Zhihong	<i>Medtronic International Ltd</i> <i>Temasek Polytechnic (BioMedical Engineering Faculty)</i> <i>Singapore Manufacturing Federation (Medical Technology Industry Group)</i> <i>Biomedical Engineering Society (Singapore)</i> <i>Eu Yan Sang International Ltd</i> <i>Health Sciences Authority</i> <i>National University of Singapore</i> <i>TüV SüD PSB Pte Ltd</i> <i>Nanyang Polytechnic</i> <i>Nanyang Technological University</i> <i>Acumen Research Laboratories Pte Ltd</i> <i>Temasek Polytechnic (Centre of Innovation for Complementary Health Products)</i> <i>Advent Access Pte Ltd</i> <i>Enterprise Singapore</i> <i>Singapore Health Services Pte Ltd</i> <i>Biosensors International Group</i> <i>Diagnostics Development (DxD) Hub</i> <i>Singapore Bioimaging Consortium</i>

The Technical Committee on Medical Devices, appointed by the Biomedical and Health Standards Committee, consists of representatives from the following organisations:

	Name	Capacity
Chairman	: Prof James Goh	<i>Individual Capacity</i>
Secretary	: Ms Iris Peng	<i>Singapore Manufacturing Federation – Standards Development Organisation</i>
Members	: Ms Chua Chui Khim	<i>Becton Dickinson Medical (S) Pte Ltd</i>
	Dr Christopher Lam	<i>Health Sciences Authority</i>
	Assoc Prof Leo Hwa Liang	<i>Biomedical Engineering Society (Singapore)</i>
	Dr Lim Jing	<i>Osteopore International Pte Ltd</i>
	Ms Iris Tan	<i>Advent Access Pte Ltd</i>

The National Mirror Working Group on IEC SC62D MT20, appointed by the Technical Committee to assist in the preparation of this standard, comprises the following experts who contribute in their individual capacity:

	Name
Convenor	: Dr Christian Gert Bluechel
Secretary	: Mr Kevin Tan
Members	: Mr Peter Haywood
	Dr Ho Teck Tuak
	Mr Watson Ong
	Mr Peh Ruey Feng
	Prof Tong Yen Wah

The organisations in which the experts of the National Mirror Working Group are involved are:

Advent Access Pte Ltd
AWAK Technologies Pte Ltd
Kidney Dialysis Foundation
National University of Singapore
Temasek Polytechnic
The Institution of Engineers, Singapore

CONTENTS

NATIONAL FOREWORD	7
FOREWORD	8
INTRODUCTION TO THE AMENDMENT	11
INTRODUCTION	12
 1 * Scope, object and related standards	 13
1.1 Scope	13
1.2 Object	13
1.3 Related standards	13
2 Normative references	14
3 Terms and definitions	14
4 General requirements	18
5 ME EQUIPMENT identification marking and documents	18
5.1 Indicator lights and controls	18
5.2 ACCOMPANYING DOCUMENTS	18
6 ALARM SYSTEMS	19
6.1 ALARM CONDITION	19
6.2 * Disclosures for INTELLIGENT ALARM SYSTEM	20
6.3 Generation of ALARM SIGNALS	20
6.4 * Disclosure of delays	26
6.5 ALARM PRESETS	27
6.6 ALARM LIMIT	29
6.7 * ALARM SYSTEM security	30
6.8 * ALARM SIGNAL inactivation states	30
6.9 * ALARM RESET	33
6.10 * NON-LATCHING and LATCHING ALARM SIGNALS	33
6.11 * DISTRIBUTED ALARM SYSTEM	33
6.12 * ALARM SYSTEM logging	35
 Annex A (informative) General guidance and rationale	 36
Annex B (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS	71
Annex C (normative) Symbols on marking	74
ANNEX D (informative) Guidance for auditory ALARM SIGNALS	80
ANNEX E (informative) Verbal ALARM SIGNALS	82
ANNEX F (normative) * Reserved melodies for ALARM SIGNALS	84
 Bibliography	 85
 Index of defined terms used in this collateral standard	 87

Figure 1 – Illustration of temporal characteristics of auditory ALARM SIGNALS	24
Figure A.1 – Graphical representation of components of ALARM SYSTEM delay.....	54
Table 1 – Determination of ALARM CONDITIONS and assignment of priorities	20
Table 2 – Characteristics of alarm indicator lights	21
Table 3 – * Characteristics of the BURST of auditory ALARM SIGNALS	23
Table 4 – * Characteristics of the PULSE of auditory ALARM SIGNALS	23
Table 5 – ALARM SIGNAL inactivation states	32
Table A.1 – Reference interpretation of Table F.1	70
Table A.2 – Reference interpretation of Table F.2	70
Table B.1 – Cross-reference of marking	71
Table B.2 – Cross-reference of ACCOMPANYING DOCUMENTS.....	72
Table B.3 – Cross-reference of instructions for use	72
Table B.4 – Cross-reference of technical description	73
Table C.1 – Graphical symbols for ALARM SYSTEMS.....	74
Table C.1 – Graphical symbols for ALARM SYSTEMS (<i>continued</i>)	75
Table C.1 – Graphical symbols for ALARM SYSTEMS (<i>continued</i>)	76
Table C.2 – Alternative ALARM SYSTEM related markings	79
Table D.1 – Attributes of perceived urgency	80
Table F.1 – * Equipment encoded auditory ALARM SIGNALS categorized by ALARM CONDITION and priority complying with Table 3 and Table 4.....	84
Table F.2 – * Auditory LOW PRIORITY ALARM SIGNAL complying with Table 3 and Table 4	84

National Foreword

This Singapore Standard was prepared by the National Mirror Working Group on IEC SC62D MT20 appointed by the Technical Committee on Medical Devices under the direction of the Biomedical and Health Standards Committee.

This standard is identical with IEC 60601-1-8:2006 + A1:2012, “Medical electrical equipment – Part 1-8: General requirements for basic safety and essential performance – Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems”, published by the International Electrotechnical Commission.

Attention is drawn to the following:

1. Where appropriate, the words “International Standard” shall be read as “Singapore Standard”.
2. Where reference to a particular part of IEC 60601 is made, the appropriate Singapore Standard (which is an identical adoption of that part of IEC 60601) shall be referenced.
3. The comma has been used throughout as a decimal marker whereas in Singapore Standards it is a practice to use a full point on the baseline as the decimal marker.

Attention is drawn to the possibility that some of the elements of this Singapore Standard may be the subject of patent rights. Enterprise Singapore shall not be held responsible for identifying any or all of such patent rights.

NOTE

1. *Singapore Standards (SSs) and Technical References (TRs) are reviewed periodically to keep abreast of technical changes, technological developments and industry practices. The changes are documented through the issue of either amendments or revisions.*
2. *An SS or TR is voluntary in nature except when it is made mandatory by a regulatory authority. It can also be cited in contracts making its application a business necessity. Users are advised to assess and determine whether the SS or TR is suitable for their intended use or purpose. If required, they should refer to the relevant professionals or experts for advice on the use of the document. Enterprise Singapore shall not be liable for any damages whether directly or indirectly suffered by anyone or any organisation as a result of the use of any SS or TR.*
3. *Compliance with a SS or TR does not exempt users from any legal obligations.*

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

**Part 1-8: General requirements for basic safety
and essential performance –**

**Collateral Standard: General requirements, tests and guidance for alarm
systems in medical electrical equipment and medical electrical systems**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of IEC 60601-1-8 consists of the second edition (2006) [documents 62A/519/CDV and 62A/537A/RVC] and its amendment 1 (2012) [documents 62A/824/FDIS and 62A/837/RVD]. It bears the edition number 2.1.

The technical content is therefore identical to the base edition and its amendment and has been prepared for user convenience. A vertical line in the margin shows where the base publication has been modified by amendment 1. Additions and deletions are displayed in red, with deletions being struck through.

International standard IEC 60601-1-8 has been prepared by IEC subcommittee 62A: Common aspects of electrical equipment used in medical practice of IEC technical committee 62: Electrical equipment in medical practice, and ISO subcommittee SC 3: Lung ventilators and related devices of ISO technical committee 121: Anaesthetic and respiratory equipment.

It is published as double logo standard.

IEC 60601-1-8 constitutes a collateral standard to IEC 60601-1: *Medical electrical equipment – Part 1: General requirements for safety and essential performance* hereafter referred to as the general standard.

This edition of IEC 60601-1-8 was revised to structurally align it with the 2005 edition of IEC 60601-1 and to implement the decision of IEC Subcommittee 62 A that the clause numbering structure of collateral standards written to IEC 60601-1:2005 would adhere to the form specified in ISO/IEC Directives, Part 2:2004. The principle technical changes are in Clause 4, which now recognizes that there is a general requirement for a risk management process in IEC 60601-1:2005.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. radiological equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. ALARM SYSTEMS).

In this collateral standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type. In addition, in Annex A text in italics indicates guidance that describes means to achieve the safety objectives of this collateral standard.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 6 includes Subclauses 6.1, 6.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 6.1, 6.2 and 6.3.1 are all subclauses of Clause 6).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (*).

A list of all parts of the IEC 60601 series, under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendments will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC or ISO publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for mandatory implementation nationally not earlier than 3 years from the date of publication.

IMPORTANT – The “colour inside” logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.

INTRODUCTION TO THE AMENDMENT

The second edition of IEC 60601-1-8 was published in 2006. Since its publication, an issue has been identified with respect to pulse and burst testing. In addition, issues have been raised by IEC/62D/MT 22, *Electromedical diagnostic and patient monitoring equipment*, during implementation of alarm system requirements in particular standards within their scope of work.

At the Brussels meeting, IEC/SC 62A accepted a proposal, based on ISO/TC 121/SC 3 Resolution Orebro 6, to develop the 1st amendment to IEC 60601-1-8:2006 to address the issues identified above. IEC/SC 62A – ISO/TC 121/SC 3 Joint Working Group 2, *Alarms*, was reactivated as a maintenance team to develop this amendment.

INTRODUCTION

MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS are increasingly used in medical practice. ALARM SIGNALS are frequently used to indicate unsatisfactory physiological PATIENT states, unsatisfactory functional states of the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM or to warn the OPERATOR of HAZARDS to the PATIENT or OPERATOR due to the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM. INFORMATION SIGNALS convey information that is independent of an ALARM CONDITION.

Surveys of healthcare personnel have indicated significant discontent with ALARM SIGNALS. Problems include difficulty in identifying the source of an ALARM SIGNAL, loud and distracting ALARM SIGNALS, and the high incidence of FALSE POSITIVE or NEGATIVE ALARM CONDITIONS [16]¹⁾. Surveys of MANUFACTURERS of medical monitors demonstrated a wide variety of DEFAULT ALARM PRESETS. The leading reason for disabling ALARM SIGNALS is the large number of ALARM SIGNALS associated with FALSE POSITIVE ALARM CONDITIONS. See also bibliography.

Safety of PATIENTS depends on the ability of the OPERATOR to correctly discern the characteristics of ALARM SIGNALS. USABILITY is an important element in the design of ALARM SIGNALS that are readily discernible without being unnecessarily distracting or disturbing. This approach is intended to rationalize the current situation, to reduce confusion by limiting proliferation of ALARM SIGNALS and their control states, and to minimize distraction for other people. This collateral standard was developed with contributions from clinicians, engineers and applied psychologists.

The terminology, requirements, general recommendations and guidance of this collateral standard are intended to be useful for MANUFACTURERS of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS and for technical committees responsible for particular standards.

The effectiveness of any ALARM SYSTEM depends critically on its implementation by the RESPONSIBLE ORGANIZATION. It is important that the RESPONSIBLE ORGANIZATION configure the ALARM SYSTEM so that an OPERATOR is not able to compromise it.

1) Figures in brackets refer to the bibliography.

MEDICAL ELECTRICAL EQUIPMENT –

**Part 1-8: General requirements for basic safety
and essential performance –**

**Collateral Standard: General requirements, tests and guidance for alarm
systems in medical electrical equipment and medical electrical systems**

1 * Scope, object and related standards

1.1 Scope

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, hereafter referred to as ME EQUIPMENT and ME SYSTEMS.

This collateral standard specifies requirements for ALARM SYSTEMS and ALARM SIGNALS in ME EQUIPMENT and ME SYSTEMS.

It also provides guidance for the application of ALARM SYSTEMS.