

TR ISO/TS 17665-2: 2018 ISO/TS 17665-2:2009, IDT

(ICS 11.080.01)

TECHNICAL REFERENCE

Sterilisation of health care products — Moist heat —

Part 2: Guidance on the application of ISO 17665-1



Published by



TR ISO/TS 17665-2 : 2018 ISO/TS 17665-2:2009, IDT (ICS 11.080.01)

TECHNICAL REFERENCE

Sterilisation of health care products — Moist heat —

Part 2: Guidance on the application of ISO 17665-1

All rights reserved. Unless otherwise specified, no part of this Technical Reference may be reproduced or utilised in any form or by any means, electronic or mechanical, including photocopying and microfilming, without permission in writing from Enterprise Singapore. Request for permission can be sent to: standards@enterprisesg.gov.sg.

© Enterprise Singapore 2018

ISBN 978-981-48-3542-8

This Technical Reference was approved on 3 December 2018 by the Biomedical and Health Standards Committee under the purview of the Singapore Standards Council.

First published, 2019

The Biomedical and Health Standards Committee, appointed by the Standards Council, consists of the following members:

		Name	Capacity
Chairman Deputy Chairmen	:	Dr Yong Chern Chet	Individual Capacity
	:	Mr Vincent Cheung Ms Selina Seah Ms Wong Woei Jiuang	Individual Capacity Changi General Hospital Health Sciences Authority
Advisor	:	Ms Jacqueline Monteiro	Individual Capacity
Secretary	:	Mr Kevin Tan	Singapore Manufacturing Federation – Standards Development Organisation
Members	:	Mr Alec Chow Boon Kuan	Medtronic International Ltd
		Mr Chung Kwong Yuew	Temasek Polytechnic (BioMedical Engineering Faculty)
		Ms Heidi Goh	Singapore Manufacturing Federation (Medical Technology Industry Group)
		Prof James Goh	Biomedical Engineering Society (Singapore)
		Dr Lai Choon Sheen	Eu Yan Sang International Ltd
		Dr Christopher Lam	Health Sciences Authority
		Assoc Prof Leo Hwa Liang	National University of Singapore
		Dr Lin Jianhua	TüV SüD PSB Pte Ltd
		Dr Leonard Loh	Nanyang Polytechnic
		Assoc Prof Eddie Ng Yin Kwee	Nanyang Technological University
		Dr Ong Siew Hwa	Acumen Research Laboratories Pte Ltd
		Dr Padmanabhan Saravanan	Temasek Polytechnic (Centre of Innovation for Complementary Health Products)
		Mr Peh Ruey Feng	Advent Access Pte Ltd
		Ms Celine Tan	Enterprise Singapore
		Prof Tan Puay Hoon	Singapore Health Services Pte Ltd
		Ms Wang Dan	Biosensors International Group
		Dr Sidney Yee	Diagnostics Development (DxD) Hub
		Dr Zhou Zhihong	Singapore Bioimaging Consortium

The Technical Committee on Quality Management Systems, appointed by the Biomedical and Health Standards Committee, consists of representatives from the following organisations:

		Name	Capacity
Chairman	:	Ms Heidi Goh	Individual Capacity
Secretary	:	Mr Kevin Tan	Singapore Manufacturing Federation – Standards Development Organisation
Members	:	Ms Jasmine Chan	Veredus Laboratories Pte Ltd
		Mr Chin Kai Hwee	Biosensors International Group
		Ms Katherine Goh	Singapore Accreditation Council
		Dr Christopher Lam	Health Sciences Authority
		Mr Nishith Desai	Medtronic International Ltd
		Ms Grace Tan	Edward Lifesciences (Singapore) Pte Ltd

The National Mirror Working Group on ISO/TC 210, appointed by the Technical Committee to assist in the preparation of this standard, comprises the following experts who contribute in their individual capacity:

Name

Convenor : Dr Margam Chandrasekaran

Secretary : Mr Kevin Tan
Members : Ms Heidi Goh

Ms How Pei Sin Ms Audrey Lee Mr Liew Ee Bin Mr Narayanan Sethu Mr Caleb Ng

Mr Paul Tan

The organisations in which the experts of the National Mirror Working Group are involved are:

Access-2-Healthcare
BioPharmaSpec UK Ltd
Sanmina Corporation Singapore
Singapore Manufacturing Federation (Medical Technology Industry Group)
SysteMED Pte Ltd
TüV SüD PSB Pte Ltd
Wise Consultants and Services Pte Ltd

(blank page)

Conte	e nts F	age
Nation	al Forewordal Foreword	6
Forew	ord	7
Introdu	uction	8
1	Scope	9
2	Normative references	9
3	Terms and definitions	10
4	Quality management system elements	10
5	Sterilizing agent characterization	10
5.1 5.2	Sterilizing agent Microbicidal effectiveness	
5.2 5.3	Material effects	
5.4	Environmental considerations	11
6	Process and equipment characterization	
6.1 6.2	Process Equipment	
7	Product definition	
8	Process definition	
9	Validation	
9.1	General	19
9.2 9.3	Installation qualification (IQ) Operational qualification (OQ)	
9.4	Performance qualification (PQ)	21
9.5	Review and approval of the validation	
10	Routine monitoring and control	
11	Product release from sterilization	25
12	Maintaining process effectiveness	
12.1 12.2	Demonstration of continued effectiveness	
12.3	Maintenance of equipment	26
12.4 12.5	RequalificationAssessment of change	
	A (informative) Evaluation of a sterilization process primarily based on the	
	measurement of physical parameters	28
Annex	B (informative) Evaluation of a sterilization process primarily based on biological inactivation and an accompanying mechanical air removal procedure	36
Annex	C (informative) Temperature and pressure of saturated steam for use in moist heat sterilization	39
Annex	D (informative) Special considerations for health care settings	41
Annex	E (informative) Index of normative clauses/subclauses of ISO 17665-1 and cited references or related guidance given in ISO 17665-1 and ISO/TS 17665-2	50
Riblion	granhy	53

National Foreword

This Technical Reference was prepared by the National Mirror Working Group on ISO/TC 210 appointed by the Technical Committee on Quality Management Systems under the direction of the Biomedical and Health Standards Committee.

This standard is identical with ISO/TS 17665-2:2009, "Sterilization of health care products — Moist heat — Part 2: Guidance on the application of ISO 17665-1", published by the International Organization for Standardization.

Attention is drawn to the following:

 The reference to International Standards shall be replaced by the following Singapore Standards:

International Standard Corresponding Singapore Standard

ISO 17665-1 SS ISO 17665-1 TR ISO/TS 17665-2

2. The comma has been used throughout as a decimal marker whereas in Singapore Standards it is a practice to use a full point on the baseline as the decimal marker.

Attention is drawn to the possibility that some of the elements of this Technical Reference may be the subject of patent rights. Enterprise Singapore shall not be held responsible for identifying any or all of such patent rights.

NOTE

- 1. Singapore Standards (SSs) and Technical References (TRs) are reviewed periodically to keep abreast of technical changes, technological developments and industry practices. The changes are documented through the issue of either amendments or revisions.
- 2. An SS or TR is voluntary in nature except when it is made mandatory by a regulatory authority. It can also be cited in contracts making its application a business necessity. Users are advised to assess and determine whether the SS or TR is suitable for their intended use or purpose. If required, they should refer to the relevant professionals or experts for advice on the use of the document. Enterprise Singapore shall not be liable for any damages whether directly or indirectly suffered by anyone or any organisation as a result of the use of any SS or TR.
- 3. Compliance with a SS or TR does not exempt users from any legal obligations

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

In other circumstances, particularly when there is an urgent market requirement for such documents, a technical committee may decide to publish other types of document:

- an ISO Publicly Available Specification (ISO/PAS) represents an agreement between technical experts in an ISO working group and is accepted for publication if it is approved by more than 50 % of the members of the parent committee casting a vote;
- an ISO Technical Specification (ISO/TS) represents an agreement between the members of a technical committee and is accepted for publication if it is approved by 2/3 of the members of the committee casting a vote.

An ISO/PAS or ISO/TS is reviewed after three years in order to decide whether it will be confirmed for a further three years, revised to become an International Standard, or withdrawn. If the ISO/PAS or ISO/TS is confirmed, it is reviewed again after a further three years, at which time it must either be transformed into an International Standard or be withdrawn.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO/TS 17665-2 was prepared by Technical Committee ISO/TC 198, Sterilization of health care products.

ISO 17665 consists of the following parts, under the general title *Sterilization of health care products* — *Moist heat*:

- Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices
- Part 2: Guidance on the application of ISO 17665-1 [Technical Specification]

Introduction

The guidance given in this Technical Specification is not intended as a checklist for assessing compliance with ISO 17665-1. This guidance is intended to assist in obtaining a uniform understanding and implementation of ISO 17665-1 by providing explanations and acceptable methods for achieving compliance with specified requirements. It highlights important aspects and provides examples. Methods other than those given in this guidance may be used. However, the use of alternative methods has to be demonstrated to be effective in achieving compliance with ISO 17665-1.

The main body of this document is applicable to all settings where moist heat sterilization is carried out. The annexes to this guidance document also specify detailed means of implementing the requirements of ISO 17665-1 and represent current best practices.

The numbering of the clauses in the main body of this Technical Specification corresponds to that in ISO 17665-1.

Medical devices reprocessed in health care facilities include a wide variety of product with varying levels of bioburden. Appropriate and thorough cleaning and, where necessary for safe handling, decontamination processes are essential prior to presenting product for sterilization. Mixed product loads are common in healthcare facilities with throughput volumes dictated by historical and predicted demand for sterile product.

Health care facilities do not normally specify sterilization processes for any individual medical device. Also, it is impractical for health care facilities to determine bioburden on a medical device. It is important that specified instruments be disassembled before decontamination and thoroughly inspected after completion of the sterilization process. Reassembly and assessment of functionality are also needed. Therefore, the medical device manufacturer's instructions (see ISO 17664^[23]) should be followed for all aspects of cleaning, disinfection, packaging and sterilization. Many devices can be fully immersed and can be washed and disinfected in automated equipment (see ISO 15883^[19-22]). For devices that cannot be fully immersed and that cannot tolerate thermal decontamination, alternative methods of disinfection should be used to ensure safe handling. Such procedures and policies should be in place to ensure that medical devices undergo appropriate reprocessing. Particular attention needs to be paid to the drying and storage of sterile medical devices. Requirements for packaging of medical devices are covered in ISO 11607-1^[8] and ISO 11607-2^[9].

If multiple sterilization cycles can lead to degradation and limit the useful life of a medical device, the manufacturer will specify the number of reprocessing cycles that can normally be tolerated.

When selecting a medical device, priority should be given to properties such as ease of cleaning and disassembly.

Additional guidance specific to health care is offered in Annex D of this Technical Specification.

Sterilization of health care products — Moist heat —

Part 2:

Guidance on the application of ISO 17665-1

1 Scope

This Technical Specification provides general guidance on the development, validation and routine control of moist heat sterilization processes and is intended to explain the requirements set forth in ISO 17665-1. The guidance given in this Technical Specification is provided to promote good practice related to moist heat sterilization processes and to assist those developing and validating a moist heat sterilization process according to ISO 17665-1.

NOTE 1 The structure of the main body of this ISO Technical Specification (Clauses 1 to 12) corresponds to the structure of ISO 17665-1, so that the guidance given under a particular clause or subclause of this part of ISO 17665 applies to the requirements given in the corresponding clause or subclause of ISO 17665-1. For example, guidance for subclause 5.2 of ISO 17665-1:2006 is given in 5.2. This guidance is provided in addition to the guidance given in ISO 17665-1:2006, Annex A. See also Annexe E.

NOTE 2 Special considerations specific to sterilization processes performed in health care facilities are given in Annex D.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 17665-1:2006, Sterilization of health care products — Moist heat — Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices

NOTE The normative references in ISO 17665-1 refer to published standards, the content of which should be used to assist in demonstrating compliance to the clause in which they are cited. Some are required mainly for moist heat sterilization in industry or for manufacturers of moist heat sterilizers and could go beyond typical practice for those performing sterilization in health care facilities.

ISO 17665-1 specifies a number of methods and procedures that can be used to monitor sterilization processes. The equipment required will normally be commercially available. A number of the normative references cited describe the specification and test methods used by commercial suppliers to qualify their products. The user of such products should ensure that purchased products comply with these standards, but will not normally need to refer to the standards.

ISO 17665-1 specifies the use of packaging complying with ISO 11607-1 and ISO 11607-2. Healthcare facilities should purchase packaging complying with these International Standards.

One method of process validation specified in ISO 17665-1 is based on the determination of bioburden. The ISO 11737^{[6],[7]} series specifies a number of microbiological methods used during this process. Health care facilities would not normally utilize this approach for process validation.