

SS IEC 60601-1-9:2018+A1:2020
IEC 60601-1-9:2007+AMD2:2020, IDT
(ICS 11.040; 13.020)

SINGAPORE STANDARD

Medical electrical equipment

– Part 1-9 : General requirements for basic safety
and essential performance – Collateral standard:
Requirements for environmentally conscious design

Incorporating Amendment No. 1

SS IEC 60601-1-9 :2018+A1:2020

IEC 60601-1-9:2007+AMD2:2020, IDT
(ICS 11.040; 13.020)

SINGAPORE STANDARD

Medical electrical equipment

- Part 1-9 : General requirements for basic safety and essential performance
- Collateral standard: Requirements for environmentally conscious design

Published by Enterprise Singapore



**THIS PUBLICATION IS COPYRIGHT
PROTECTED**

Copyright © 2020 Enterprise Singapore
Copyright © 2020 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilised in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either Enterprise Singapore, representing the IEC National Committee of Singapore, or the IEC. If you have any questions about the copyrights of Enterprise Singapore or the IEC or have an enquiry about obtaining additional rights to this publication, please contact Enterprise Singapore at: standards@enterprisesg.gov.sg for further information.

ISBN 978-981-49-2572-3

The content of this Singapore Standard was approved on 15 November 2018 by the Biomedical and Health Standards Committee (BHSC) under the purview of the Singapore Standards Council.

First published, 2019

BHSC consists of the following members:

	Name	Representation
Chairman	: Dr Yong Chern Chet	<i>Individual Capacity</i>
Deputy Chairmen	: Mr Vincent Cheung	<i>Individual Capacity</i>
	Ms Selina Seah	<i>Changi General Hospital</i>
	Ms Wong Woei Jiuang	<i>Health Sciences Authority</i>
Advisor	: Ms Jacqueline Monteiro	<i>Individual Capacity</i>
Secretary	: Mr Kevin Tan	<i>Singapore Manufacturing Federation – Standards Development Organisation</i>
Members	: Mr Alec Chow Boon Kuan	<i>Medtronic International Ltd</i>
	Mr Chung Kwong Yuew	<i>Temasek Polytechnic (BioMedical Engineering Faculty)</i>
	Ms Heidi Goh	<i>Singapore Manufacturing Federation (Medical Technology Industry Group)</i>
	Prof James Goh	<i>Biomedical Engineering Society (Singapore)</i>
	Dr Lai Choon Sheen	<i>Eu Yan Sang International Ltd</i>
	Dr Christopher Lam	<i>Health Sciences Authority</i>
	Assoc Prof Leo Hwa Liang	<i>National University of Singapore</i>
	Dr Lin Jianhua	<i>TÜV SÜD PSB Pte Ltd</i>
	Dr Leonard Loh	<i>Nanyang Polytechnic</i>
	Assoc Prof Eddie Ng Yin Kwee	<i>Nanyang Technological University</i>
	Dr Ong Siew Hwa	<i>Acumen Research Laboratories Pte Ltd</i>
	Dr Padmanabhan Saravanan	<i>Temasek Polytechnic (Centre of Innovation for Complementary Health Products)</i>
	Mr Peh Ruey Feng	<i>Advent Access Pte Ltd</i>
	Ms Celine Tan	<i>Enterprise Singapore</i>
	Prof Tan Puay Hoon	<i>Singapore Health Services Pte Ltd</i>
	Ms Wang Dan	<i>Biosensors International Group</i>
	Dr Sidney Yee	<i>Diagnostics Development (DxD) Hub</i>
	Dr Zhou Zhihong	<i>Singapore Bioimaging Consortium</i>

BHSC set up the Technical Committee on Medical Devices to oversee the preparation of this standard. The Technical Committee consists of the following members:

	Name	Representation
Chairman	: Prof James Goh	<i>Individual Capacity</i>
Secretary	: Mr Kevin Tan	<i>Singapore Manufacturing Federation – Standards Development Organisation</i>
Members	: Ms Chua Chui Khim	<i>Becton Dickinson Medical (S) Pte Ltd</i>
	Dr Christopher Lam	<i>Health Sciences Authority</i>
	Assoc Prof Leo Hwa Liang	<i>Biomedical Engineering Society (Singapore)</i>
	Dr Lim Jing	<i>Osteopore International Pte Ltd</i>
	Ms Iris Tan	<i>Advent Access Pte Ltd</i>

The Technical Committee set up the National Mirror Working Group on IEC SC62D MT20 to prepare this standard. The Working Group consists of the following experts who contribute in their *individual capacity*:

	Name
Convenor	: Dr Christian Gert Bluechel
Secretary	: Mr Kevin Tan
Members	: Mr Peter Haywood
	Dr Ho Teck Tuak
	Mr Watson Ong
	Mr Peh Ruey Feng
	Prof Tong Yen Wah

The organisations in which the experts of the Working Group are involved are:

Advent Access Pte Ltd
AWAK Technologies Pte Ltd
Kidney Dialysis Foundation
National University of Singapore
Temasek Polytechnic
The Institution of Engineers, Singapore

(blank page)

CONTENTS

NATIONAL FOREWORD	6
FOREWORD	7
INTRODUCTION	10
INTRODUCTION to Amendment 1	11
INTRODUCTION to Amendment 2	11
1 Scope, object and related standards	12
1.1 * Scope	12
1.2 Object	12
1.3 Related standards	12
2 Normative references	13
3 Terms and definitions	13
4 Protection of the ENVIRONMENT	15
4.1 * Identification of ENVIRONMENTAL ASPECTS	15
4.2 * Determination of significant ENVIRONMENTAL ASPECTS	15
4.3 * Information from the SUPPLY CHAIN	16
4.4 * Reduction of adverse ENVIRONMENTAL IMPACTS	16
4.5 Environmental information	16
Annex A (informative) General guidance and rationale	18
Annex B (informative) Guide to marking and labelling requirements for ME EQUIPMENT	33
Bibliography	34
Index of defined terms used in this collateral standard	35
Table A.1 – Example product LIFE-CYCLE stages	19
Table A.2 – Examples of ENVIRONMENTAL IMPACTS and their cause	25
Table A.3 – ENVIRONMENTAL ASPECTS and typical ENVIRONMENTAL IMPACTS	28
Table B.1 – ACCOMPANYING DOCUMENTS, General	33
Table B.2 – Other information	33

National Foreword

This Singapore Standard was prepared by the National Mirror Working Group on IEC SC62D MT20 set up by the Technical Committee on Medical Devices under the purview of BHSC.

This standard is identical with IEC 60601-1-9:2007, “Medical electrical equipment – Part 1-9: General requirements for basic safety and essential performance – Collateral standard: Requirements for environmentally conscious design”, including the amendments to this edition, published by the International Electrotechnical Commission.

NOTE 1 – Where appropriate, the words “International Standard” are read as “Singapore Standard”.

NOTE 2 – Reference to International/Overseas Standards are replaced by applicable Singapore Standards or Technical References.

Attention is drawn to the possibility that some of the elements of this Singapore Standard may be the subject of patent rights. Enterprise Singapore shall not be held responsible for identifying any or all of such patent rights.

NOTE

1. *Singapore Standards (SSs) and Technical References (TRs) are reviewed periodically to keep abreast of technical changes, technological developments and industry practices. The changes are documented through the issue of either amendments or revisions. Where SSs are deemed to be stable, i.e. no foreseeable changes in them, they will be classified as “Mature Standards”. Mature Standards will not be subject to further review, unless there are requests to review such standards.*
2. *An SS or TR is voluntary in nature except when it is made mandatory by a regulatory authority. It can also be cited in contracts making its application a business necessity. Users are advised to assess and determine whether the SS or TR is suitable for their intended use or purpose. If required, they should refer to the relevant professionals or experts for advice on the use of the document. Enterprise Singapore and the Singapore Standards Council shall not be liable for any damages whether directly or indirectly suffered by anyone or any organisation as a result of the use of any SS or TR. Although care has been taken to draft this standard, users are also advised to ensure that they apply the information after due diligence.*
3. *Compliance with a SS or TR does not exempt users from any legal obligations.*

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of the official IEC Standard and its amendments has been prepared for user convenience.

IEC 60601-1-9 edition 1.2 contains the first edition (2007-07) [documents 62A/571/FDIS and 62A/575/RVD], its amendment 1 (2013-06) [documents 62A/874/FDIS and 62A/881/RVD] and its amendment 2 (2020-07) [documents 62A/1393/FDIS and 62A/1408/RVD].

This Final version does not show where the technical content is modified by amendments 1 and 2. A separate Redline version with all changes highlighted is available in this publication.

International standard IEC 60601-1-9 has been prepared by IEC subcommittee 62A: Common aspects of electrical equipment used in medical practice, of IEC technical committee 62: Electrical equipment in medical practice.

The first edition of this publication constitutes a collateral standard to IEC 60601-1: *Medical electrical equipment – Part 1: General requirements for safety and essential performance* hereafter referred to as the general standard.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. radiological equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. alarm systems).

In this collateral standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the four numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 4 includes subclauses 4.1, 4.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 4.1, 4.5 and 4.5.1 are all subclauses of Clause 4).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (*).

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendments will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

NOTE The attention of National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC or ISO publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for mandatory implementation nationally not earlier than 3 years from the date of publication.

INTRODUCTION

The objective of this collateral standard is to improve the ENVIRONMENTAL IMPACT for the entire range of MEDICAL ELECTRICAL EQUIPMENT, taking into account all stages of the product LIFE CYCLE:

- product specification;
- design;
- manufacturing;
- sales, logistics, installation;
- use;
- END OF LIFE management.

This means protecting the ENVIRONMENT and human health from HAZARDOUS SUBSTANCES, conserving raw materials and energy, minimizing the generation of WASTE, as well as minimizing the adverse ENVIRONMENTAL IMPACTS associated with WASTE. The criteria needed to reach this goal must be integrated into all stages of the MEDICAL ELECTRICAL EQUIPMENT LIFE CYCLE from the specification stage to END OF LIFE management.

The ENVIRONMENTAL IMPACTS of ME EQUIPMENT through all LIFE-CYCLE stages are determined from the MEDICAL ELECTRICAL EQUIPMENT'S ENVIRONMENTAL ASPECTS defined during the identification of need, product planning, and design stages (see Table A.1). Consideration of ENVIRONMENTAL ASPECTS as early as possible in these stages can produce numerous benefits that might include lower costs, stimulation of innovation and creativity, and increased knowledge about the product. It can also provide new business opportunities, and improved product quality as well as reduction of adverse ENVIRONMENTAL IMPACTS. The assessment of the ENVIRONMENTAL ASPECTS and IMPACTS of MEDICAL ELECTRICAL EQUIPMENT is a developing science and it is anticipated that this collateral standard will require periodic updating as the science develops.

The requirements given in this collateral standard do not replace national or international laws and regulations.

Environmental protection is one element of the overall RISK MANAGEMENT PROCESS as required by the general standard.

The acceptability of MEDICAL ELECTRICAL EQUIPMENT'S ENVIRONMENTAL IMPACTS are balanced against other factors, such as the product's intended function, performance, safety, cost, marketability, quality, legal and regulatory requirements. This balance can differ depending on the intended function of the MEDICAL ELECTRICAL EQUIPMENT. For example, a solution appropriate for life-saving or life-supporting MEDICAL ELECTRICAL EQUIPMENT might not be appropriate for a device intended to correct a minor ailment. A MANUFACTURER of MEDICAL ELECTRICAL EQUIPMENT might have to justify, as a result of RISK MANAGEMENT, that a medical benefit outweighs the associated adverse ENVIRONMENTAL IMPACTS.

INTRODUCTION to Amendment 1

The first edition of IEC 60601-1-9 was published in 2007. This amendment is intended to update the references to IEC 60601-1:2005 to include Amendment 1:2012 and to make a few minor editorial updates.

INTRODUCTION to Amendment 2

The first edition of IEC 60601-1-9 was published in 2007 and amended in 2013. Since the publication of IEC 60601-1-9:2007+A1:2013, the IEC Subcommittee (SC) 62A Secretariat has been collecting issues from a variety of sources including comments from National Committees. At the November 2015 meeting of IEC/SC 62A in Kobe, Japan, the subcommittee initiated a process to identify high-priority issues that need to be considered in an amendment and should not wait until the second edition of IEC 60601-1-9, which is presently targeted for publication sometime after 2024.

As directed in item 1 of Kobe Resolution 1, the IEC/SC 62A Chairman Advisory Group (CAG) considered the 7 issues collected by the SC/62A Secretariat for IEC 60601-1-9:2007 and determined that none met the selection criteria stated in Kobe Resolution 1.

However, an amendment is needed to update the reference to IEC 60601-1:2005+A1:2012 +A2:2020. In London in 2018, SC 62A approved the development of an administrative amendment to IEC 60601-1-9:2007+A1:2013.

Because this is an amendment to IEC 60601-1-9:2007, the style in force at the time of publication of IEC 60601-1-9 has been applied to this amendment. The specified in ISO/IEC Directives Part 2:2018 has only been applied when implementing the new style guidance would not result in additional editorial changes.

Users of this document should note that when constructing the dated references to specific elements in a standard, such as definitions, amendments are only referenced if they modified the text being cited. For example, if a reference is made to a definition that has not been modified by an amendment, then the reference to the amendment is not included in the dated reference.

MEDICAL ELECTRICAL EQUIPMENT –

Part 1-9: General requirements for basic safety and essential performance – Collateral Standard: Requirements for environmentally conscious design

Scope, object and related standards

1.1 * Scope

This International Standard applies to the reduction of adverse ENVIRONMENTAL IMPACTS of MEDICAL ELECTRICAL EQUIPMENT, hereafter referred to as ME EQUIPMENT.

MEDICAL ELECTRICAL SYSTEMS are excluded from the scope of this collateral standard.

1.2 Object

The object of this collateral standard is to specify general requirements, in addition to those of the general standard, for the reduction of the adverse ENVIRONMENTAL IMPACT of ME EQUIPMENT, and to serve as the basis for particular standards.

1.3 Related standards

1.3.1 IEC 60601-1

For ME EQUIPMENT, this collateral standard complements IEC 60601-1.

When referring to IEC 60601-1 or to this collateral standard, either individually or in combination, the following conventions are used:

- "the general standard" designates IEC 60601-1 alone, including any amendments;
- "this collateral standard" designates IEC 60601-1-9 alone, including any amendments;
- "this standard" designates the combination of the general standard and this collateral standard.

1.3.2 Particular standards

A requirement in a particular standard takes priority over the corresponding requirement in this collateral standard.

1.3.3 Environmental standards

This standard takes into account the ISO 14000 series of environmental standards with particular emphasis on ISO 14062 [8]¹⁾.

¹⁾ Figures in square brackets refer to the Bibliography.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

Amendment 1:2012

Amendment 2:2020