

SS IEC 60601-1-12:2018+A1:2020
IEC 60601-1-12:2014+AMD1:2020, IDT
(ICS 11.040)

SINGAPORE STANDARD

Medical electrical equipment

– Part 1-12 : General requirements for basic safety and essential performance – Collateral standard:
Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment

Incorporating Amendment No. 1

SS IEC 60601-1-12:2018+A1:2020
IEC 60601-1-12:2014+AMD1:2020, IDT
(ICS 11.040)

SINGAPORE STANDARD

Medical electrical equipment

- Part 1-12: General requirements for basic safety and essential performance
- Collateral standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment

Published by Enterprise Singapore



**THIS PUBLICATION IS COPYRIGHT
PROTECTED**
Copyright © 2020 Enterprise Singapore
Copyright © 2020 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilised in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either Enterprise Singapore, representing the IEC National Committee of Singapore, or the IEC. If you have any questions about the copyrights of Enterprise Singapore or the IEC or have an enquiry about obtaining additional rights to this publication, please contact Enterprise Singapore at: standards@enterprisesg.gov.sg for further information.

ISBN 978-981-49-2576-1

The content of this Singapore Standard was approved on 15 November 2018 by the Biomedical and Health Standards Committee (BHSC) under the purview of the Singapore Standards Council.

First published, 2019

BHSC consists of the following members:

	Name	Representation
Chairman	: Dr Yong Chern Chet	<i>Individual Capacity</i>
Deputy Chairmen	: Mr Vincent Cheung	<i>Individual Capacity</i>
	: Ms Selina Seah	<i>Changi General Hospital</i>
	: Ms Wong Woei Jiuang	<i>Health Sciences Authority</i>
Advisor	: Ms Jacqueline Monteiro	<i>Individual Capacity</i>
Secretary	: Ms Iris Peng	<i>Singapore Manufacturing Federation – Standards Development Organisation</i>
Members	: Mr Alec Chow Boon Kuan	<i>Medtronic International Ltd</i>
	Mr Chung Kwong Yuew	<i>Temasek Polytechnic (BioMedical Engineering Faculty)</i>
	Ms Heidi Goh	<i>Singapore Manufacturing Federation (Medical Technology Industry Group)</i>
	Prof James Goh	<i>Biomedical Engineering Society (Singapore)</i>
	Dr Lai Choon Sheen	<i>Eu Yan Sang International Ltd</i>
	Dr Christopher Lam	<i>Health Sciences Authority</i>
	Assoc Prof Leo Hwa Liang	<i>National University of Singapore</i>
	Dr Lin Jianhua	<i>TÜV SÜD PSB Pte Ltd</i>
	Dr Leonard Loh	<i>Nanyang Polytechnic</i>
	Assoc Prof Eddie Ng Yin Kwee	<i>Nanyang Technological University</i>
	Dr Ong Siew Hwa	<i>Acumen Research Laboratories Pte Ltd</i>
	Dr Padmanabhan Saravanan	<i>Temasek Polytechnic (Centre of Innovation for Complementary Health Products)</i>
	Mr Peh Ruey Feng	<i>Advent Access Pte Ltd</i>
	Ms Celine Tan	<i>Enterprise Singapore</i>
	Prof Tan Puay Hoon	<i>Singapore Health Services Pte Ltd</i>
	Ms Wang Dan	<i>Biosensors International Group</i>
	Dr Sidney Yee	<i>Diagnostics Development (DxD) Hub</i>
	Dr Zhou Zhihong	<i>Singapore Bioimaging Consortium</i>

BHSC set up the Technical Committee on Medical Devices to oversee the preparation of this standard. The Technical Committee consists of the following members:

	Name	Representation
Chairman	: Prof James Goh	<i>Individual Capacity</i>
Secretary	: Ms Iris Peng	<i>Singapore Manufacturing Federation – Standards Development Organisation</i>
Members	: Ms Chua Chui Khim	<i>Becton Dickinson Medical (S) Pte Ltd</i>
	Dr Christopher Lam	<i>Health Sciences Authority</i>
	Assoc Prof Leo Hwa Liang	<i>Biomedical Engineering Society (Singapore)</i>
	Dr Lim Jing	<i>Osteopore International Pte Ltd</i>
	Ms Iris Tan	<i>Advent Access Pte Ltd</i>

The Technical Committee set up the National Mirror Working Group on IEC SC62D MT20 to prepare this standard. The Working Group consists of the following experts who contribute in their *individual capacity*:

	Name
Convenor	: Dr Christian Gert Bluechel
Secretary	: Mr Kevin Tan
Members	: Mr Peter Haywood
	Dr Ho Teck Tuak
	Mr Watson Ong
	Mr Peh Ruey Feng
	Prof Tong Yen Wah

The organisations in which the experts of the Working Group are involved are:

Advent Access Pte Ltd
AWAK Technologies Pte Ltd
Kidney Dialysis Foundation
National University of Singapore
Temasek Polytechnic
The Institution of Engineers, Singapore

(blank page)

CONTENTS

NATIONAL FOREWORD	7
FOREWORD	8
INTRODUCTION	11
INTRODUCTION to Amendment 1	11
1 Scope, object and related standards	12
1.1 * Scope	12
1.2 * Object	12
1.3 Related standards	13
1.3.1 IEC 60601-1	13
1.3.2 Particular standards	13
2 Normative references	13
3 Terms and definitions	14
4 General requirements.....	15
4.1 * Additional requirements for SUPPLY MAINS for ME EQUIPMENT and ME SYSTEMS	15
4.2 * Environmental conditions for ME EQUIPMENT.....	15
4.2.1 * Environmental conditions of transport and storage between uses	16
4.2.2 * Environmental operating conditions	17
5 * Classification of ME EQUIPMENT and ME SYSTEMS.....	20
6 ME EQUIPMENT identification, marking and documents.....	20
6.1 * Additional requirements for legibility of markings	20
6.2 * Additional requirements for marking of IP classification	20
6.3 * Instructions for use	21
6.3.1 Additional general requirements	21
6.3.2 * Additional requirements for an electrical power source.....	21
6.3.3 Additional requirements for ME EQUIPMENT start-up PROCEDURE.....	22
6.3.4 * Additional requirements for operating instructions.....	22
6.3.5 Additional requirements for ME EQUIPMENT messages	22
6.4 Technical description – FIXED or PERMANENTLY INSTALLED CLASS I ME EQUIPMENT	23
7 * Protection against electrical HAZARDS from ME EQUIPMENT	23
8 Protection against excessive temperatures and other HAZARDS.....	23
8.1 Additional requirements for ingress of water or particulate matter into ME EQUIPMENT and ME SYSTEMS	23
8.1.1 * Ingress of water or particulate matter into ME EQUIPMENT.....	23
8.1.2 * Ingress of water or particulate matter into ME SYSTEMS.....	23
8.2 Additional requirements for interruption of the power supply to ME EQUIPMENT and ME SYSTEM.....	24
8.3 * Additional requirements for INTERNAL ELECTRICAL POWER SOURCE for ME EQUIPMENT	25
9 * Accuracy of controls and instruments and protection against hazardous outputs.....	25
10 Construction of ME EQUIPMENT	26

10.1	* Additional requirements for mechanical strength of ME EQUIPMENT intended for the EMS ENVIRONMENT	26
10.1.1	General requirements for mechanical strength	26
10.1.2	* Requirements for mechanical strength for FIXED or PERMANENTLY INSTALLED ME EQUIPMENT intended for use in a road ambulance	26
10.1.3	* Requirements for mechanical strength for TRANSPORTABLE ME EQUIPMENT	27
10.1.4	* Requirements for mechanical strength for ME EQUIPMENT intended for airborne use	28
10.2	Requirements for mounting of ME EQUIPMENT	29
11	Additional requirements for electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	29
Annex A (informative)	General guidance and rationale	31
A.1	General guidance	31
A.2	Rationale for particular clauses and subclauses	33
Annex B (informative)	Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS	47
B.1	Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts	47
B.2	ACCOMPANYING DOCUMENTS, instructions for use	47
B.3	ACCOMPANYING DOCUMENTS, technical description	48
Annex C (informative)	Symbols on marking	49
Bibliography	51
Index of defined terms used in this collateral standard.....		53
Figure A.1 – Saturation water vapour pressure as function of temperature		36
Table 1 – Mechanical strength test applicability		26
Table A.1 – Saturation water vapour pressure as function of temperature.....		37
Table B.1 – Marking on the outside of ME EQUIPMENT, ME SYSTEMS or their parts.....		47
Table B.2 – ACCOMPANYING DOCUMENTS, instructions for use.....		47
Table B.3 – ACCOMPANYING DOCUMENTS, technical description		48
Table C.1 – General symbols		49

National Foreword

This Singapore Standard was prepared by the National Mirror Working Group on IEC SC62D MT20 set up by the Technical Committee on Medical Devices under the purview of BHSC.

This standard is identical with IEC 60601-1-12:2014, "Medical electrical equipment – Part 1-12: General requirements for basic safety and essential performance – Collateral standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment", including the amendments to this edition, published by the International Electrotechnical Commission.

NOTE 1 – Where appropriate, the words "International Standard" are read as "Singapore Standard".

NOTE 2 – Reference to International Standards are replaced by applicable Singapore Standards/Technical References.

NOTE 3 – Where numerical values are expressed as decimals, the comma is read as a full point.

Attention is drawn to the possibility that some of the elements of this Singapore Standard may be the subject of patent rights. Enterprise Singapore shall not be held responsible for identifying any or all of such patent rights.

NOTE

- 1. Singapore Standards (SSs) and Technical References (TRs) are reviewed periodically to keep abreast of technical changes, technological developments and industry practices. The changes are documented through the issue of either amendments or revisions. Where SSs are deemed to be stable, i.e. no foreseeable changes in them, they will be classified as "Mature Standards". Mature Standards will not be subject to further review, unless there are requests to review such standards.*
- 2. An SS or TR is voluntary in nature except when it is made mandatory by a regulatory authority. It can also be cited in contracts making its application a business necessity. Users are advised to assess and determine whether the SS or TR is suitable for their intended use or purpose. If required, they should refer to the relevant professionals or experts for advice on the use of the document. Enterprise Singapore and the Singapore Standards Council shall not be liable for any damages whether directly or indirectly suffered by anyone or any organisation as a result of the use of any SS or TR. Although care has been taken to draft this standard, users are also advised to ensure that they apply the information after due diligence.*
- 3. Compliance with a SS or TR does not exempt users from any legal obligations.*

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This consolidated version of the official IEC Standard and its amendment has been prepared for user convenience.

IEC 60601-1-12 edition 1.1 contains the first edition (2014-06) [documents 62A/932/FDIS and 62A/938/RVD] and its amendment 1 (2020-07) [documents 62A/1396/FDIS and 62A/1411/RVD].

This Final version does not show where the technical content is modified by amendment 1. A separate Redline version with all changes highlighted is available in this publication.

International standard IEC 60601-1-12 has been prepared by a joint working group of IEC subcommittee 62A: Common aspects of electrical equipment used in medical practice, of IEC technical committee 62: Electrical equipment in medical practice, and ISO subcommittee SC3: Lung ventilators and related devices, of ISO technical committee 121: Anaesthetic and respiratory equipment.

This first edition constitutes a collateral standard to IEC 60601-1 (third edition): *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance* hereafter referred to as the general standard.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. radiological equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. ALARM SYSTEMS).

In this collateral standard, the following print types are used:

- requirements and definitions: roman type.
- *test specifications: italic type.*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.3.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this collateral standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (*).

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendment will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of Member Bodies and National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised ISO or IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

Medical practice is increasingly using MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS for monitoring, treatment or diagnosis of PATIENTS in the EMERGENCY MEDICAL SERVICES ENVIRONMENT (see 3.1). The safety of MEDICAL ELECTRICAL EQUIPMENT in this uncontrolled, rough environment is a cause for concern.

This collateral standard was developed with contributions from clinicians, engineers and regulators. The terminology, requirements, general recommendations and guidance of this collateral standard are intended to be useful for MANUFACTURERS of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS and for technical committees responsible for the development of particular standards.

INTRODUCTION to Amendment 1

The first edition of IEC 60601-1-12 was published in 2014. Since the publication of IEC 60601-1-12:2014, the IEC Subcommittee (SC) 62A Secretariat has been collecting issues from a variety of sources including comments from National Committees. At the November 2015 meeting of IEC/SC 62A in Kobe, Japan, the subcommittee initiated a process to identify high-priority issues that need to be considered in an amendment and should not wait until the second edition of IEC 60601-1-12, which is presently targeted for publication sometime after 2024.

As directed in item 1 of Kobe Resolution 1, the IEC/SC 62A Chairman Advisory Group (CAG) considered the 27 issues collected by the SC/62A Secretariat for IEC 60601-1-12:2014 and determined that none met the selection criteria stated in Kobe Resolution 1.

However, an amendment is needed to update the references to IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020. In London in 2018, SC 62A approved the development of an administrative amendment to IEC 60601-1-12:2014.

Because this is an amendment to IEC 60601-1-12:2014, the style in force at the time of publication of IEC 60601-1-12 has been applied to this amendment. The style specified in ISO/IEC Directives Part 2:2018 has only been applied when implementing the new style guidance would not result in additional editorial changes.

Users of this document should note that when constructing the dated references to specific elements in a standard, such as definitions, amendments are only referenced if they modified the text being cited. For example, if a reference is made to a definition that has not been modified by an amendment, then the reference to the amendment is not included in the dated reference.

MEDICAL ELECTRICAL EQUIPMENT –

Part 1-12: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment

1 Scope, object and related standards

1.1 * Scope

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, hereafter referred to as ME EQUIPMENT and ME SYSTEMS, which are intended, as indicated in the instructions for use by their MANUFACTURER, for use in the EMS ENVIRONMENT (EMERGENCY MEDICAL SERVICES ENVIRONMENT), as defined in 3.1.

NOTE 1 For the purposes of this standard, the intent of the MANUFACTURER is indicated in the instructions for use. The RESPONSIBLE ORGANIZATION and the OPERATOR need to be aware that any other use outside the MANUFACTURER'S INTENDED USE can result in a HAZARDOUS SITUATION for the PATIENT.

The EMS ENVIRONMENT includes

- responding to and providing life support at the scene of an emergency to a PATIENT reported as experiencing injury or illness in a pre-hospital setting, and transporting the PATIENT, while continuing such life support care, to an appropriate professional healthcare facility for further care.
- providing monitoring, treatment or diagnosis during transport between professional healthcare facilities.

This International Standard does not apply to ME EQUIPMENT and ME SYSTEMS intended solely for use in the HOME HEALTHCARE ENVIRONMENT covered by IEC 60601-1-11 or solely for use in professional healthcare facilities covered by IEC 60601-1 without the additions of IEC 60601-1-11 or this collateral standard. ME EQUIPMENT and ME SYSTEMS are often not solely intended for one environment. Such ME EQUIPMENT or ME SYSTEM can be intended for multiple use environments, and as such, if also intended for use in the EMS ENVIRONMENT, are within the scope of this standard.

EXAMPLE ME EQUIPMENT or ME SYSTEM intended for both the EMS ENVIRONMENT and the professional healthcare facility environment.

NOTE 2 EMS ENVIRONMENT ME EQUIPMENT and ME SYSTEMS can be used in locations with unreliable electrical sources and outdoor environmental conditions.

1.2 * Object

The object of this collateral standard is to provide general requirements for ME EQUIPMENT and ME SYSTEMS carried to the scene of an emergency and used there, as well as in transport, in situations where the ambient conditions differ from indoor conditions.

The object of this collateral standard is to specify general requirements that are in addition to those of the general standard and to serve as the basis for particular standards.

1.3 Related standards

1.3.1 IEC 60601-1

For ME EQUIPMENT and ME SYSTEMS, this collateral standard complements IEC 60601-1.

When referring to IEC 60601-1 or to this collateral standard, either individually or in combination, the following conventions are used:

- "the general standard" designates IEC 60601-1 alone, including any amendments;
- "this collateral standard" designates IEC 60601-1-12 alone, including any amendments;
- "this standard" designates the combination of the general standard and this collateral standard.

1.3.2 Particular standards

A requirement in a particular standard takes priority over the corresponding requirement in this collateral standard.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

NOTE 2 Informative references are listed in the bibliography on page 51.

IEC 60068-2-27:2008, *Environmental testing – Part 2-27: Tests – Test Ea and guidance: Shock*

IEC 60068-2-31:2008, *Environmental testing – Part 2-31: Tests – Test Ec: Rough handling shocks, primarily for equipment-type specimens*

IEC 60068-2-64:2008, *Environmental testing – Part 2-64: Tests – Test Fh: Vibration, broadband random and guidance*

IEC 60529:1989, *Degrees of protection provided by enclosures (IP Code)*
IEC 60529:1989/AMD1:1999¹

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1:2005/AMD1:2012

IEC 60601-1:2005/AMD2:2020

IEC 60601-1-2:2014, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests*

IEC 60601-1-2:2014/AMD1:2020

¹ There exists a consolidated edition 2.1(2001) including IEC 60529:1989 and its Amendment 1:1999.

IEC 60601-1-6:2010, *Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability*
IEC 60601-1-6:2010/AMD1:2013
IEC 60601-1-6:2010/AMD2:2020

IEC 60601-1-8:2006, *Medical electrical equipment – Part 1-8: General requirements for basic safety and essential performance – Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems*
IEC 60601-1-8:2006/AMD1:2012
IEC 60601-1-8:2006/AMD1:2020

IEC 60601-1-11:2015, *Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*
IEC 60601-1-11:2015/AMD1:2020

CISPR 11:2009, *Industrial, scientific and medical equipment – Radio-frequency disturbance characteristics – Limits and methods of measurement*

ISO 7000:2014, *Graphical symbols for use on equipment – Registered symbols*

ISO 7010:2019, *Graphical symbols – Safety colours and safety signs – Registered safety signs*

ISO 15223-1:2012, *Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements*

EUROCAE² ED-14G, *Environmental conditions and test procedures for airborne equipment*

RTCA³ DO-160G, *Environmental Conditions and Test Procedures for Airborne Equipment*

2 EUROCAE (European Organization for Civil Aviation Electronics), 102 rue Etienne Dolet, 92240 Malakoff, France.

3 RTCA (Radio Technical Commission for Aeronautics), 1150 18th St, NW., Suite 910, Washington, DC 20036, USA.